Table Of Contents

ANALYSIS: VOA METHOD NO.: 8260

ARDL Report(s) #: 006613 MATRIX: LIMU (SEA WEED)

	PAGE
Organic Package Review Checklist	2
Case Narrative and/or Nonconformance Reports	5
Sample Receipt Information	8
Field Sample Results	13
QC Summary Forms	22
Sample Preparation Information	35
Standard Preparation Information	36
Analytical Run Data	70
Example Calculation	179
Miscellaneous Additional Data	182

	By:	Date:
Package Preparation	JAA	03/12/14
Analytical Review	LS	3/24/14
Final Review	BM	4/02/14
QA Review	DSS	4/2/14

Rev. 1.1

Organic Package Review Checklist

Rev. 1.1

ORGANIC PACKAGE REVIEW CHECKLIST

METHOD: 8260

	Initial	Date
Case Narrative and/or Nonconformance Reports		
Case Narrative / Analyst Comment Sheets	US	3/24/14
Nonconformance / Noncompliance Reports (if applicable)		
Sample Receipt Information		
Copies of Original Chain-of-Custody Forms	US	3/24/14
Cooler Receipt Form and Airbills		
Phone Messages (If applicable)		
Field Sample Results		
QC Summary Forms		
Surrogate Summary, if applicable	US	3/24/14
Spike Blank and/or MS Summary		
Blank Results & Blank Summary, if applicable		
ISTD Area Summary, if applicable		
Sample Preparation Information (if applicable)		
Total Solids Determination (If applicable)	Nen	
Extraction Log (If applicable)		
Extraction Notebook Pages (If applicable)		
Spiking Solutions Notebook Pages		
MS/MSD's		
Surrogate/ISTDs		
Spiking Solutions Certificates Of Analyses		
MS/MSD's		
Surrogates/ISTDs		
Certificates of Analysis (Vendor)		
Standard Preparation Information		
Instrument Logs/Analysis Notebook Pages	US	3/24/14
Certificates of Analysis (Vendor)		
Analytical Run Number 1		
Calibration Information	US	3/24/14
Tune Data		
Initial Calibration		
Continuing Calibration		
Instrument Logs/Analysis Notebook Pages		
Blank/Sample/Field QC Data		
Analytical Run Number 2 (if required)		
Calibration Information	US	3/24/14
Tune Data		
Initial Calibration		
Continuing Calibration		
Instrument Logs/Analysis Notebook Pages		
Blank/Sample/Field QC Data		

Rev. 1.1

ORGANIC PACKAGE REVIEW CHECKLIST (Continued)

METHOD: 8260

	Initial	Date
Analytical Run Number 3 (if required)		
Calibration Information	LS	3/24/14
Tune Data		
Initial Calibration		
Continuing Calibration		
Instrument Logs/Analysis Notebook Pages		
Blank/Sample/Field QC Data		
Analytical Run Number 4 (if required)		
Calibration Information	Na	
Tune Data		
Initial Calibration		
Continuing Calibration		
Instrument Logs/Analysis Notebook Pages		
Blank/Sample/Field QC Data		
Analytical Run Number 5 (if required)		
Calibration Information	Na	
Tune Data		
Initial Calibration		
Continuing Calibration		
Instrument Logs/Analysis Notebook Pages		
Blank/Sample/Field QC Data		
Analytical Run Number 6 (if required)		
Calibration Information	Na	
Tune Data		
Initial Calibration		
Continuing Calibration		
Instrument Logs/Analysis Notebook Pages		
Blank/Sample/Field QC Data		
Example Calculation		
Example Calculation Sheet	LS	3/24/14
Miscellaneous Additional Data (if applicable)	LS	3/24/14

NA = Not Applicable

I have checked this package for completeness and verify the following:

1. Pencils, whiteout, whiteout tape, and dark highlighters were not used. ✓
2. All notebook pages are signed. ✓
3. All corrections contain the original data "struck through" with initials and date. ✓
4. All comments found throughout the data are defined as required. ✓
5. All pages are legible. ✓
6. Manual integrations, if applicable, are printed out and include an explanation with signature of the analyst and date. Na

Package Completeness Checked By/Date: LS 3/24/14

Rev. 1.1

Case Narrative and/or Nonconformance Reports

Rev. 1.1

ORGANIC ANALYSIS CASE NARRATIVE

Corps of Engineers – ERDC, Vicksburg **Date: 4/02/14**

ARDL Report No.: 6613 **Lab Name: ARDL, Inc.**

Project Name: MAKAU MILITARY RESERVATION OAHU (LIMU)

Samples Received at ARDL: 02/20/2014

Lab ID No.	Sample ID No.	Date Collected	Date Analyzed	Analysis Requested
006613-01	MAK101L	01/13/2014	02/28/2014	Purgeable Organics
006613-02	MAK102L	01/13/2014	02/28/2014	Purgeable Organics
006613-03	MAK105L	01/13/2014	02/28/2014	Purgeable Organics
006613-04	MAK106L	01/14/2014	02/27/2014	Purgeable Organics
006613-05	MAK108, 109, 110, 113L	01/14/2014	02/27/2014	Purgeable Organics
006613-06	MAK111L	01/14/2014	02/27/2014	Purgeable Organics
006613-07	MAK112L	01/14/2014	02/28/2014	Purgeable Organics
006613-08	MAK115L	01/14/2014	02/28/2014	Purgeable Organics

PURGEABLE ORGANICS – METHOD 8260B

Eight (8) tissue (LIMU) samples were received by ARDL, Inc. on February 20, 2014, for 8260B analysis. Due to specific project requirements and the unusual nature of the samples, some changes to standard protocol were made. The requested analyte list was limited to ethylbenzene, m & p-xylene, o-xylene, toluene, styrene and 1,2,4-trimethylbenzene. Only the two internal standards used to quantitate these analytes and their associated surrogate compounds (toluene-d8 and 4-bromofluorobenzene) were evaluated in all calibrations and sample analyses. A nominal sample weight of 2 grams was chosen to avoid problems related to the sample matrix while achieving a reporting limit of 12.5 ug/Kg. It should be noted that all limu samples required analysis at 1 gram in order to pass internal standard response criteria. Reporting limits have been adjusted accordingly. Due to the tendency of these samples to foam during the purge step, a commercial anti-foam agent (Restek Cat. No.31822) was added to all samples, blanks and LCS analyses. A massive, co-eluting interferent peak at approximately the 9 minute retention time was present in all the LIMU samples. This compound was determined to be bromoform, a compound reported in the literature to be produced by seaweed. Since the bromoform mass spectra has significant M/Z 91 ion, quantitative interference with nearby o-xylene was experienced. Consequently, all calibrations and sample analyses for o-xylene used the secondary characteristic ion M/Z 106 for quantitation.

INITIAL CALIBRATION

The initial calibration was performed using eight calibration points and passed criteria.

CONTINUING CALIBRATION

All daily calibration standards passed criteria.

SURROGATES

All surrogate recoveries met the default range of 50 – 150%.

INTERNAL STANDARDS

Internal standard responses were acceptable for all analyses.

ORGANIC ANALYSIS CASE NARRATIVE

Corps of Engineers – ERDC, Vicksburg **Date: 4/02/14**

ARDL Report No.: 6613 **Lab Name: ARDL, Inc.**

Project Name: MAKAU MILITARY RESERVATION OAHU (LIMU)

Samples Received at ARDL: 02/20/2014

LCS

All LCS analyses met the recovery criteria.

MATRIX SPIKE/MATRIX SPIKE DUPLICATE

An MS/MSD analyses was performed on sample MAK101L and passed recovery criteria except for styrene and o-Xylene as noted on summary forms. The depressed styrene and o-Xylene recoveries are attributed to the background bromoform interference.

METHOD BLANKS

No target analytes were detected in the method blanks above the reporting limit.

ORGANIC DATA REPORTING QUALIFIERS

The following organic data reporting qualifiers are used as required.

ND- Indicates compound was analyzed for but not detected. The sample quantitation limit must be corrected for dilution and for percent moisture.

J - Indicates an estimated value. This flag is used when estimating a concentration of a compound that meets identification criteria but the result is less than the quantitation or reporting limit but greater than the method detection limit.

P - This flag is used for any two column confirmation methodology when there is greater than 40% difference for detected concentrations between the two columns. The primary column value is reported on the Form 1 and flagged with a "P".

B - This flag is used when the analyte is found above the reporting limit in the blank as well as the sample. It indicates possible/probable blank contamination and warns the data user to take appropriate action.

Release of the data contained in this package has been authorized by the Technical Services Manager or his designee as verified by the following signature.

Dean S. Dickerson
Technical Services Manager

Sample Receipt Information

Rev. 1.1

ARDL, Inc

P.O. Box 1566, 400 Aviation Drive, Mount Vernon, Illinois 62864
Phone (618) 244-3235, Fax (618) 244-1149

CHAIN OF CUSTODY RECORD

Project
ARDL / Makau Military Reservation Oahu, Hawaii Project

SAMPLERS (Signature)

SAMPLE NUMBER	DATE	TIME	Comp	Grab	# of Containers	Purgeable Organics 8260B	Explosives 8330	OCPs 8081	SVCOs 8270C									ARDL, Inc Remarks or Sample Location	PRESERVATION ICED	PRESERVATION Specify Chemicals added and final pH if known
MAK101O	1/13/2014	10:02		X	1	X	x	x	x									Octopus	x	
MAK102O	1/13/2014	10:30		X	1	X	x	x	x									Octopus	x	
MAK103O	1/13/2014	11:02		X	1	X	x	x	x									Octopus	x	
MAK104O	1/13/2014	11:30		X	1	X	x	x	x									Octopus	x	
MAK105O	1/13/2014	11:58		X	1	X	x	x	x									Octopus	x	
MAK106O	1/13/2014	12:20		X	1	X	x	x	x									Octopus	x	
MAK107O	1/13/2014	13:05		X	1	X	x	x	x									Octopus	x	
MAK108O	1/13/2014	14:20		X	1	X	x	x	x									Octopus	x	
MAK102C	1/13/2014	10:02		X	1	X	x	x	x									Sea Cucumber	x	
MAK103C	1/13/2014	10:22		X	1	X	x	x	x									Sea Cucumber	x	
MAK104C	1/13/2014	11:00		X	1	X	x	x	x									Sea Cucumber	x	
MAK105C	1/13/2014	11:30		X	1	X	x	x	x									Sea Cucumber	x	
MAK106C	1/13/2014	12:05		X	1	X	x	x	x									Sea Cucumber	x	
MAK107C	1/13/2014	12:40		X	1	X	x	x	x									Sea Cucumber	x	
MAK108C	1/13/2014	13:50		X	1	X	x	x	x									Sea Cucumber	x	
MAK110C	1/14/2014	9:55		X	1	X	x	x	x									Sea Cucumber	x	
MAK101L	1/13/2014	9:10		X	1	X	x	x	x									Limu (seaweed)	x	
MAK102L	1/13/2014	10:02		X	1	X	x	x	x									Limu (seaweed)	x	
MAK105L	1/13/2014	15:20		X	1	X	x	x	x									Limu (seaweed)	x	
MAK106L	1/14/2014	8:45		X	1	X	x	x	x									Limu (seaweed)	x	
MAK108,109,110,113L	1/14/2014		X		1	X	x	x	x									Limu (seaweed)	x	
MAK111L	1/14/2014	12:05		X	1	X	x	x	x									Limu (seaweed)	x	
MAK112L	1/14/2014	12:50		X	1	X	x	x	x									Limu (seaweed)	x	
MAK115L	1/14/2014	14:00		X	1	X	x	x	x									Limu (seaweed)	x	

Relinquished by: (Signature)	Date	Time	Relinquished by: (Signature)
PACE	2/19/14	1600	
Relinquished by: (Signature)	Date	Time	Relinquished by: (Signature)
Received for Laboratory by: (Signature)	Date	Time	Shipping Ticket No.
L Cochrum	2/20/14	0920	

REMARKS / SPECIAL INSTRUCTIONS:

Please send level 4 hardcopy report and excel edd to: Dean Dickerson, c/o ARDL, Inc. 400 Aviation Drive, Mount Vernon, IL 62864. Email questons to ddickerson@ardlinc.com.

COOLER RECEIPT REPORT
ARDL, INC.

ARDL #: 6611, 6612, 6613

Cooler # HH001

Number of Coolers in Shipment: 1

Project: MAKAU Military Reservation Oahu, Hawaii Project

Date Received: 2-20-14

A. **PRELIMINARY EXAMINATION PHASE:** Date cooler was opened: 2-20-14 (Signature) D L Cackrum

1. Did cooler come with a shipping slip (airbill, etc.)? ..... (YES) NO

   If YES, enter carrier name and airbill number here: fedex 9325 1888 0829

2. Were custody seals on outside of cooler? ..... YES (NO) N/A

   How many and where? _____, Seal Date: _____, Seal Name: _____

3. Were custody seals unbroken and intact at the date and time of arrival? ..... YES NO (N/A)
4. Did you screen samples for radioactivity using a Geiger Counter? ..... (YES) NO
5. Were custody papers sealed in a plastic bag ~~and taped inside to the lid?~~ dlc 2/20/14 ..... (YES) NO
6. Were custody papers filled out properly (ink, signed, etc.)? ..... (YES) NO N/A
7. Were custody papers signed in appropriate place by ARDL personnel? ..... (YES) NO N/A
8. Was project identifiable from custody papers? If YES, enter project name at the top of this form ..... (YES) NO N/A
9. Was a separate container provided for measuring temperature? YES ____ NO ✓ Cooler Temp. -6.5 C

B. **LOG-IN PHASE:** Date samples were logged-in: 2-20-14 (Signature) D L Cackrum

10. Describe type of packing in cooler: blue ice / bubble bags
11. Were all bottles sealed in separate plastic bags? in bags / not sealed ..... YES NO N/A
12. Did all bottles arrive unbroken and were labels in good condition? ..... YES (NO)
13. Were bottle labels complete? ..... (YES) NO
14. Did all bottle labels agree with custody papers? ..... (YES) NO
15. Were correct containers used for the tests indicated? ..... (YES) NO
16. Was pH correct on preserved water samples? ..... YES NO (N/A)
17. Was a sufficient amount of sample sent for tests indicated? ..... (YES) NO
18. Were bubbles absent in VOA samples? If NO, list by sample #: _____ YES NO (N/A)
19. Was the ARDL project coordinator notified of any deficiencies? ..... YES NO (N/A)

Comments and/or Corrective Action:
12) MAK 115L Cracked lid / seal intact
(By: Signature) dlc Date: 2-20-14

Sample Transfer	
Fraction: All	Fraction:
Area #: Freezer	Area #:
By: dlc	By:
On: 2-20-14	On:

M:\ADMIN\FORMS\COOLER RECEIPT REPORT.doc Rev. 05/07/07

ORIGIN ID:GRBA (920) 468-2436
SHIPPING
PACE ANALYTICAL
1241 BELLEVUE

GREEN BAY, WI 54302
UNITED STATES US

SHIP DATE: 19FEB14
ACTWGT: 64.0 LB
CAD: 255289/CAFE2704

DRY ICE: 4.5 KGS
BILL THIRD PARTY

TO DEAN DICKERSON
ARDL
400 AVIATION DRIVE
P.O BOX 1566
MOUNT VERNON IL 62864
(618) 244-3235
DEPT: 1301-9501

518C1/562F/6F03

FedEx Express

J13111305230126

THU - 20 FEB AA
STANDARD OVERNIGHT

TRK# 0201 9325 1888 0829

ICE
62864
IL-US STL

- NA MVNA

Part # 156148-434 RIT2 03/12

ORIGIN ID:GRBA (920) 469-2436
SHIPPING
PACE ANALYTICAL
1241 BELLEVUE

GREEN BAY, WI 54302
UNITED STATES US

SHIP DATE: 19FEB14
ACTWGT: 64.0 LB
CAD: 255289/CAFE2704

DRY ICE: 4.5 KGS
BILL THIRD PARTY

TO DEAN DICKERSON
ARDL
400 AVIATION DRIVE
P.O BOX 1566
MOUNT VERNON IL 62864
(618) 244-3235
DEPT: 1301-9501

518C1/56CF/6F03

FedEx Express

J131113052230126

TRK# 0201 9325 1888 0829

THU - 20 FEB AA
STANDARD OVERNIGHT
ICE

NA MVNA

62864
IL-US STL

Part # 156148-434 RIT2 03/12

Field Sample Results

Rev. 1.1

ARDL, INC.
400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	MAK101L	ARDL Lab No.:	006613-01
Desc/Location:	MAKAU MILITARY RESERV OAHU	Lab Filename:	A0228407
Sample Date:	01/13/2014	Received Date:	02/20/2014
Sample Time:	0910	Prep. Date:	02/28/2014
Matrix:	TISSUE	Analysis Date:	02/28/2014
Amount Used:	1 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSX
% Moisture:	No Moisture Present	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units	Dilution Factor
Toluene	5.0	25.0	ND		UG/KG	1
Ethyl Benzene	5.0	25.0	ND		UG/KG	1
m & p-Xylene	10.0	50.0	ND		UG/KG	1
Styrene	5.0	25.0	ND		UG/KG	1
o-Xylene	5.0	25.0	ND		UG/KG	1
1,2,4-Trimethylbenzene	5.0	25.0	ND		UG/KG	1

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	101%
4-Bromofluorobenzene	50-150	131%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Sample 006613-01, VOLATILES, GC/MS (8260) Page 1 of 1

ARDL, INC.
400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	MAK102L	ARDL Lab No.:	006613-02
Desc/Location:	MAKAU MILITARY RESERV OAHU	Lab Filename:	A0228408
Sample Date:	01/13/2014	Received Date:	02/20/2014
Sample Time:	1002	Prep. Date:	02/28/2014
Matrix:	TISSUE	Analysis Date:	02/28/2014
Amount Used:	1 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSX
% Moisture:	No Moisture Present	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units	Dilution Factor
Toluene	5.0	25.0	ND		UG/KG	1
Ethyl Benzene	5.0	25.0	ND		UG/KG	1
m & p-Xylene	10.0	50.0	ND		UG/KG	1
Styrene	5.0	25.0	ND		UG/KG	1
o-Xylene	5.0	25.0	ND		UG/KG	1
1,2,4-Trimethylbenzene	5.0	25.0	ND		UG/KG	1

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	98%
4-Bromofluorobenzene	50-150	116%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Sample 006613-02, VOLATILES, GC/MS (8260) Page 1 of 1

ARDL, INC.
400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	MAK105L	ARDL Lab No.:	006613-03
Desc/Location:	MAKAU MILITARY RESERV OAHU	Lab Filename:	A0228409
Sample Date:	01/13/2014	Received Date:	02/20/2014
Sample Time:	1520	Prep. Date:	02/28/2014
Matrix:	TISSUE	Analysis Date:	02/28/2014
Amount Used:	1 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSX
% Moisture:	No Moisture Present	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units	Dilution Factor
Toluene	5.0	25.0	ND		UG/KG	1
Ethyl Benzene	5.0	25.0	ND		UG/KG	1
m & p-Xylene	10.0	50.0	ND		UG/KG	1
Styrene	5.0	25.0	ND		UG/KG	1
o-Xylene	5.0	25.0	ND		UG/KG	1
1,2,4-Trimethylbenzene	5.0	25.0	ND		UG/KG	1

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	97%
4-Bromofluorobenzene	50-150	119%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Sample 006613-03, VOLATILES, GC/MS (8260) Page 1 of 1

ARDL, INC.
400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	MAK106L	ARDL Lab No.:	006613-04
Desc/Location:	MAKAU MILITARY RESERV OAHU	Lab Filename:	A0227413
Sample Date:	01/14/2014	Received Date:	02/20/2014
Sample Time:	0845	Prep. Date:	02/27/2014
Matrix:	TISSUE	Analysis Date:	02/27/2014
Amount Used:	1 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSW
% Moisture:	No Moisture Present	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units	Dilution Factor
Toluene	5.0	25.0	ND		UG/KG	1
Ethyl Benzene	5.0	25.0	ND		UG/KG	1
m & p-Xylene	10.0	50.0	ND		UG/KG	1
Styrene	5.0	25.0	ND		UG/KG	1
o-Xylene	5.0	25.0	ND		UG/KG	1
1,2,4-Trimethylbenzene	5.0	25.0	ND		UG/KG	1

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	97%
4-Bromofluorobenzene	50-150	133%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

ARDL, INC.
400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	MAK108,109,110,113L	ARDL Lab No.:	006613-05
Desc/Location:	MAKAU MILITARY RESERV OAHU	Lab Filename:	A0227414
Sample Date:	01/14/2014	Received Date:	02/20/2014
Sample Time:	---	Prep. Date:	02/27/2014
Matrix:	TISSUE	Analysis Date:	02/27/2014
Amount Used:	1 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSW
% Moisture:	No Moisture Present	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units	Dilution Factor
Toluene	5.0	25.0	ND		UG/KG	1
Ethyl Benzene	5.0	25.0	ND		UG/KG	1
m & p-Xylene	10.0	50.0	ND		UG/KG	1
Styrene	5.0	25.0	ND		UG/KG	1
o-Xylene	5.0	25.0	ND		UG/KG	1
1,2,4-Trimethylbenzene	5.0	25.0	ND		UG/KG	1

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	98%
4-Bromofluorobenzene	50-150	125%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Sample 006613-05, VOLATILES, GC/MS (8260) Page 1 of 1

ARDL, INC.
400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	MAK111L	ARDL Lab No.:	006613-06
Desc/Location:	MAKAU MILITARY RESERV OAHU	Lab Filename:	A0227415
Sample Date:	01/14/2014	Received Date:	02/20/2014
Sample Time:	1205	Prep. Date:	02/27/2014
Matrix:	TISSUE	Analysis Date:	02/27/2014
Amount Used:	1 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSW
% Moisture:	No Moisture Present	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units	Dilution Factor
Toluene	5.0	25.0	ND		UG/KG	1
Ethyl Benzene	5.0	25.0	ND		UG/KG	1
m & p-Xylene	10.0	50.0	ND		UG/KG	1
Styrene	5.0	25.0	ND		UG/KG	1
o-Xylene	5.0	25.0	ND		UG/KG	1
1,2,4-Trimethylbenzene	5.0	25.0	ND		UG/KG	1

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	102%
4-Bromofluorobenzene	50-150	133%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Sample 006613-06, VOLATILES, GC/MS (8260) Page 1 of 1

ARDL, INC.
400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	MAK112L	ARDL Lab No.:	006613-07
Desc/Location:	MAKAU MILITARY RESERV OAHU	Lab Filename:	A0228410
Sample Date:	01/14/2014	Received Date:	02/20/2014
Sample Time:	1250	Prep. Date:	02/28/2014
Matrix:	TISSUE	Analysis Date:	02/28/2014
Amount Used:	1 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSX
% Moisture:	No Moisture Present	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units	Dilution Factor
Toluene	5.0	25.0	ND		UG/KG	1
Ethyl Benzene	5.0	25.0	ND		UG/KG	1
m & p-Xylene	10.0	50.0	ND		UG/KG	1
Styrene	5.0	25.0	ND		UG/KG	1
o-Xylene	5.0	25.0	ND		UG/KG	1
1,2,4-Trimethylbenzene	5.0	25.0	ND		UG/KG	1

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	98%
4-Bromofluorobenzene	50-150	129%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

ARDL, INC.
400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	MAK115L	ARDL Lab No.:	006613-08
Desc/Location:	MAKAU MILITARY RESERV OAHU	Lab Filename:	A0228411
Sample Date:	01/14/2014	Received Date:	02/20/2014
Sample Time:	1400	Prep. Date:	02/28/2014
Matrix:	TISSUE	Analysis Date:	02/28/2014
Amount Used:	1 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSX
% Moisture:	No Moisture Present	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units	Dilution Factor
Toluene	5.0	25.0	ND		UG/KG	1
Ethyl Benzene	5.0	25.0	ND		UG/KG	1
m & p-Xylene	10.0	50.0	ND		UG/KG	1
Styrene	5.0	25.0	ND		UG/KG	1
o-Xylene	5.0	25.0	ND		UG/KG	1
1,2,4-Trimethylbenzene	5.0	25.0	ND		UG/KG	1

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	99%
4-Bromofluorobenzene	50-150	124%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Sample 006613-08, VOLATILES, GC/MS (8260) Page 1 of 1

QC Summary Forms

Surrogate Summary, if applicable

Spike Blank and/or MS Summary

Blank Results & Blank Summary, if applicable

ISTD Area Summary, if applicable

Rev. 1.1

2B

SEA WEED VOLATILE SURROGATE RECOVERY

Lab Name: ARDL, INC. Contract: ---

Lab Code: --- Case No.: 006613 SAS No.: --- SDG No.: ---

	EPA SAMPLE NO.	S1 (TOL) #	S2 (BFB) #	S3 ( ) #	S4 ( ) #			TOT OUT
01	SPBK 02/27/14	100	98					0
02	BLANK 02/27/14	101	111					0
03	MAK106L	97	133					0
04	MAK108,109,110,113L	98	125					0
05	MAK111L	102	133					0
06	SPBK 02/28/14	102	114					0
07	BLANK 02/28/14	99	116					0
08	MAK101L	101	131					0
09	MAK102L	98	116					0
10	MAK105L	97	119					0
11	MAK112L	98	129					0
12	MAK115L	99	124					0
13	SPBK 03/03/14	115	116					0
14	BLANK 03/03/14	107	111					0
15	MAK101LMS	97	108					0
16	MAK101LMSD	110	114					0
17								
18								
19								
20								
21								
22								
23								
24								
25								
26								
27								
28								
29								
30								

		QC LIMITS	MATRIX LIMITS
S1	Toluene-d8	(50 - 150)	(50 - 150)
S2	4-Bromofluorobenzene	(50 - 150)	(50 - 150)
S3		(-)	(-)
S4		(-)	(-)

column to be used to flag recovery values
* Values outside of contract required QC limits
D Surrogates diluted out

BLANK SPIKE/SPIKE DUPLICATE REPORT

ARDL, INC. 400 Aviation Drive; P.O. Box 1566 Mt. Vernon, IL 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260) Analytical Method: 8260B
Project No.: Prep Method: 5030A

Matrix: QC Material QC Batch: 0305JFSW Prep. Date: 02/27/2014
Amount Used: 2 g Level: LOW Analysis Date: 02/27/2014

Parameter	Spike Result	Spike Level	Spike % Rec	Duplicate Result	Duplicate Level	Duplicate % Rec	Recovery Limits	RPD	RPD Limit
Toluene	96.1	100	96	--	--	--	85-118	--	--
Ethyl Benzene	97.1	100	97	--	--	--	86-119	--	--
m & p-Xylene	197	200	99	--	--	--	81-121	--	--
Styrene	96.2	100	96	--	--	--	84-119	--	--
o-Xylene	94.2	100	94	--	--	--	81-121	--	--
1,2,4-Trimethylbenzene	102	100	102	--	--	--	83-119	--	--

SURROGATE RECOVERIES:	Spike %R	Duplicate %R	%R Limits
Toluene-d8	100	--	50-150
4-Bromofluorobenzene	97.9	--	50-150

'*' indicates a recovery outside of standard limits.
Spike Blanks for 006613-04, VOLATILES, GC/MS (8260)

BLANK SPIKE/SPIKE DUPLICATE REPORT

ARDL, INC. 400 Aviation Drive; P.O. Box 1566 Mt. Vernon, IL 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260) Analytical Method: 8260B
Project No.: Prep Method: 5030A

Matrix: QC Material QC Batch: 0305JFSX Prep. Date: 02/28/2014
Amount Used: 2 g Level: LOW Analysis Date: 02/28/2014

Parameter	Spike Result	Spike Level	Spike % Rec	Duplicate Result	Duplicate Level	Duplicate % Rec	Recovery Limits	RPD	RPD Limit
Toluene	95	100	95	--	--	--	85-118	--	--
Ethyl Benzene	93.5	100	93	--	--	--	86-119	--	--
m & p-Xylene	185	200	92	--	--	--	81-121	--	--
Styrene	93.9	100	94	--	--	--	84-119	--	--
o-Xylene	92.8	100	93	--	--	--	81-121	--	--
1,2,4-Trimethylbenzene	111	100	111	--	--	--	83-119	--	--

SURROGATE RECOVERIES:	Spike %R	Duplicate %R	%R Limits
Toluene-d8	101.7	--	50-150
4-Bromofluorobenzene	113.9	--	50-150

'*' indicates a recovery outside of standard limits.
Spike Blanks for 006613-01, VOLATILES, GC/MS (8260)

MATRIX SPIKE/SPIKE DUPLICATE REPORT

ARDL, INC. 400 Aviation Drive; P.O. Box 1566 Mt. Vernon, IL 62864

Lab Report No: 006613 Report Date: 03/21/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260) Analytical Method: 8260B
Project No.: Prep Method: 5030A

Field ID:	MAK101L	Prep. Date:	03/03/2014	ARDL Lab No.:	006613-01
Desc/Location:	MAKAU MILITARY RESERV OAH	Amount Used:	1 g	Lab Filename:	
Sample Date:	01/13/2014	% Moisture:	No Moisture Present	Received Date:	02/20/2014
Sample Time:	0910	QC Batch:	0305JFSX	Analysis Date:	03/03/2014
Matrix:	TISSUE	Level:	LOW		

Parameter	Sample Result	MS Result	MS Level	MS % Rec	MSD Result	MSD Level	MSD % Rec	% Rec Limits	RPD	RPD Limit
Toluene	ND	183	200	91.6	189	200	94.3	81-126	2.9	25
Ethyl Benzene	ND	173	200	86.4	174	200	86.8	82-130	0.5	25
m & p-Xylene	ND	340	400	84.9	339	400	84.8	84-129	0.1	25
Styrene	ND	40.1	200	20.1 *	36.5	200	18.2 *	79-127	9.5	25
o-Xylene	ND	129	200	64.5 *	122	200	60.8 *	81-128	5.8	25
1,2,4-Trimethylbenzene	ND	179	200	89.6	177	200	88.6	80-124	1.1	25

SURROGATE RECOVERIES:	MS %R	MSD %R	%R Limits
Toluene-d8	97	110	50-150
4-Bromofluorobenzene	108	114	50-150

'nc' indicates sample >4X spike level.
'*' indicates a recovery outside of standard limits.
Matrix Spikes for 006613-01, VOLATILES, GC/MS (8260)

4B

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

BLANK 02/27/14

Lab Name: ARDL, INC. Contract: ---

Lab Code: --- Case No.: 006613 SAS No.: --- SDG No.: ---

Lab File ID: A0227406 Lab Sample ID: 006613-04B1

Date Analyzed: 02/27/14 Time Analyzed: 10:53

GC Column: RTX-VMS ID: 0.25 (mm) Heated Purge: (Y/N) Y

Instrument ID: AG1

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS, AND MSD:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	TIME ANALYZED
01	MAK106L	006613-04	A0227413	14:29
02	MAK108,109,110,113L	006613-05	A0227414	14:57
03	MAK111L	006613-06	A0227415	15:25
04				
05				
06				
07				
08				
09				
10				
11				
12				
13				
14				
15				
16				
17				
18				
19				
20				
21				
22				
23				
24				
25				
26				
27				
28				
29				
30				

COMMENTS:

METHOD BLANK REPORT
ARDL, Inc. 400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	NA	ARDL Lab No.:	006613-04B1
Desc/Location:	NA	Lab Filename:	A0227406
Sample Date:	NA	Received Date:	NA
Sample Time:	NA	Prep. Date:	02/27/2014
Matrix:	QC Material	Analysis Date:	02/27/2014
Amount Used:	2 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSW
% Moisture:	NA	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units
Toluene	2.50	12.5	ND		UG/KG
Ethyl Benzene	2.50	12.5	ND		UG/KG
m & p-Xylene	5.00	25.0	ND		UG/KG
Styrene	2.50	12.5	ND		UG/KG
o-Xylene	2.50	12.5	ND		UG/KG
1,2,4-Trimethylbenzene	2.50	12.5	ND		UG/KG

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	101%
4-Bromofluorobenzene	50-150	111%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Blank for Run 0305JFSW, VOLATILES, GC/MS (8260) Page 1 of 1

4B

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

BLANK 02/28/14

Lab Name: ARDL, INC. Contract: ---

Lab Code: --- Case No.: 006613 SAS No.: --- SDG No.: ---

Lab File ID: A0228406 Lab Sample ID: 006613-01B1

Date Analyzed: 02/28/14 Time Analyzed: 11:51

GC Column: RTX-VMS ID: 0.25 (mm) Heated Purge: (Y/N) Y

Instrument ID: AG1

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS, AND MSD:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	TIME ANALYZED
01	MAK101L	006613-01	A0228407	12:25
02	MAK102L	006613-02	A0228408	13:03
03	MAK105L	006613-03	A0228409	13:44
04	MAK112L	006613-07	A0228410	14:22
05	MAK115L	006613-08	A0228411	14:50
06				
07				
08				
09				
10				
11				
12				
13				
14				
15				
16				
17				
18				
19				
20				
21				
22				
23				
24				
25				
26				
27				
28				
29				
30				

COMMENTS:

METHOD BLANK REPORT

ARDL, Inc. 400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	NA	ARDL Lab No.:	006613-01B1
Desc/Location:	NA	Lab Filename:	A0228406
Sample Date:	NA	Received Date:	NA
Sample Time:	NA	Prep. Date:	02/28/2014
Matrix:	QC Material	Analysis Date:	02/28/2014
Amount Used:	2 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSX
% Moisture:	NA	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units
Toluene	2.50	12.5	ND		UG/KG
Ethyl Benzene	2.50	12.5	ND		UG/KG
m & p-Xylene	5.00	25.0	ND		UG/KG
Styrene	2.50	12.5	ND		UG/KG
o-Xylene	2.50	12.5	ND		UG/KG
1,2,4-Trimethylbenzene	2.50	12.5	ND		UG/KG

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	99%
4-Bromofluorobenzene	50-150	116%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Blank for Run 0305JFSX, VOLATILES, GC/MS (8260) Page 1 of 1

4B

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

BLANK 03/03/14

Lab Name: ARDL, INC. Contract: ---

Lab Code: --- Case No.: 006613 SAS No.: --- SDG No.: ---

Lab File ID: A0303406 Lab Sample ID: 006613-01B2

Date Analyzed: 03/03/14 Time Analyzed: 12:15

GC Column: RTX-VMS ID: 0.25 (mm) Heated Purge: (Y/N) Y

Instrument ID: AG1

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS, AND MSD:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	TIME ANALYZED
01	MAK101LMS	006613-01MS	A0303407	12:50
02	MAK101LMSD	006613-01MD	A0303408	13:26
03				
04				
05				
06				
07				
08				
09				
10				
11				
12				
13				
14				
15				
16				
17				
18				
19				
20				
21				
22				
23				
24				
25				
26				
27				
28				
29				
30				

COMMENTS:

8A

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: ARDL, Inc. Contract: ---

Lab Code: --- Case No.: 006613 SAS No.: --- SDG No.: ---

Lab File ID (Standard): A0212416 Date Analyzed: 02/12/14

Instrument ID: AG1 Time Analyzed: 18:32

Matrix: (soil/water) SEA WEED Level: (low/med) Low Column: (pack/cap) CAP

		IS1(CBZ) AREA #	RT #	IS2(DCB) AREA #	RT #	IS3( ) AREA #	RT #
	12 HOUR STANDARD	279384	8.76	268227	10.14		
	UPPER LIMIT	558768	9.26	536454	10.64		
	LOWER LIMIT	139692	8.26	134114	9.64		
	EPA SAMPLE NO.						
1	SPBK 02/27/14	251112	8.77	270267	10.14		
2	BLANK 02/27/14	226740	8.77	245868	10.14		
3	MAK106L	198739	8.77	138647	10.14		
4	MAK108,109,110,113	203750	8.76	151097	10.14		
5	MAK111L	197159	8.76	144682	10.14		
6							
7							
8							
9							
10							
11							
12							
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							

IS1 (CBZ) = Chlorobenzene-d5
IS2 (DCB) = 1,4-Dichlorobenzene-d4
IS3 ( ) =

UPPER LIMIT = +100%
of internal standard area
LOWER LIMIT = - 50%
of internal standard area

Column used to flag internal standard area values with an asterisk
* Values outside of QC limits

8A

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: ARDL, Inc. Contract: ---

Lab Code: --- Case No.: 006613 SAS No.: --- SDG No.: ---

Lab File ID (Standard): A0212416 Date Analyzed: 02/12/14

Instrument ID: AG1 Time Analyzed: 18:32

Matrix: (soil/water) SEA WEED Level: (low/med) Low Column: (pack/cap) CAP

		IS1(CBZ) AREA #	RT #	IS2(DCB) AREA #	RT #	IS3( ) AREA #	RT #
	12 HOUR STANDARD	279384	8.76	268227	10.14		
	UPPER LIMIT	558768	9.26	536454	10.64		
	LOWER LIMIT	139692	8.26	134114	9.64		
	EPA SAMPLE NO.						
1	SPBK 02/28/14	241311	8.76	207659	10.14		
2	BLANK 02/28/14	244246	8.76	233307	10.14		
3	MAK101L	234011	8.77	145054	10.14		
4	MAK102L	255914	8.76	191516	10.14		
5	MAK105L	248663	8.76	177554	10.14		
6	MAK112L	226731	8.76	157537	10.14		
7	MAK115L	252826	8.76	185155	10.14		
8							
9							
10							
11							
12							
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							

IS1 (CBZ) = Chlorobenzene-d5
IS2 (DCB) = 1,4-Dichlorobenzene-d4
IS3 ( ) =

UPPER LIMIT = + 100%
of internal standard area
LOWER LIMIT = - 50%
of internal standard area

Column used to flag internal standard area values with an asterisk
* Values outside of QC limits

8A

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: ARDL, Inc. Contract: ---

Lab Code: --- Case No.: 006613 SAS No.: --- SDG No.: ---

Lab File ID (Standard): A0212416 Date Analyzed: 02/12/14

Instrument ID: AG1 Time Analyzed: 18:32

Matrix: (soil/water) SEA WEED Level: (low/med) Low Column: (pack/cap) CAP

		IS1(CBZ) AREA #	RT #	IS2(DCB) AREA #	RT #	IS3( ) AREA #	RT #
	12 HOUR STANDARD	279384	8.76	268227	10.14		
	UPPER LIMIT	558768	9.26	536454	10.64		
	LOWER LIMIT	139692	8.26	134114	9.64		
	EPA SAMPLE NO.						
1	SPBK 03/03/14	270704	8.77	257087	10.14		
2	BLANK 03/03/14	284953	8.77	273520	10.14		
3	MAK101LMS	349346	8.77	252292	10.15		
4	MAK101LMSD	313113	8.77	260288	10.15		
5							
6							
7							
8							
9							
10							
11							
12							
13							
14							
15							
16							
17							
18							
19							
20							
21							
22							

IS1 (CBZ) = Chlorobenzene-d5
IS2 (DCB) = 1,4-Dichlorobenzene-d4
( ) =

UPPER LIMIT = + 100%
of internal standard area
LOWER LIMIT = - 50%
of internal standard area

Column used to flag internal standard area values with an asterisk
* Values outside of QC limits

Sample Preparation Information

(Not Applicable for VOA Purge and Trap Method)

Rev. 1.1

Standards Preparation Information

Rev. 1.1

X1004 Gases - Ultra

100 ul

DWM-544-1
Lot: CJ-2970
Exp: 08/31/2015
ULTRA 1 mL
VOC Gas Mixture
6 analyte(s) at 2000 µg/mL in methanol R4318
250 Smith St, No Kingstown, RI 02852 USA
For Laboratory Use Only
US 1/29/14

3900 ul MeOH (R4474)
4000 ul 50 ug/ml US 1/29/14 Exp. 7/05/14

X1005 Acetone (only) STD

90 µL of ResTek Acetone (5000 ppm) 30245 Lot A088080 exp 07/2015
1710 µL of MeOH R4474

1,800 µL of 250 µg/mL
US 1/30/14 exp. 2/6/14

X1006 Ketones (Absolute)

150 µL Abs Ketones (2,000 ppm) Cat 82402, Lot 070313 Xp 07/2016
1050 µL MeOH

1,200 µL of 250 µg/mL US 1/30/14

X1007 BFB

50 ul STS-110N-1 lot CH-3248 2000 ug/ml (R4540) exp. 11/30/15
3950 ul MeOH (R4479)
4000 ul 25 ug/ml US 2/04/14 Exp. 8/04/14

US 2/04/14

Continued on Page

Read and Understood By

US 2/04/14 DRC 2/5/14
Signed Date Signed Date

X1021 524.2 IC MIX

43 ml 8260 Mix (W) DWM-589N-1 lot CK-2561 2000 ug/ml (R4501) exp. 9/30/16
86 ml Hexachloroethane (Abs) 70199 lot 058212 1000 ug/ml (R4497) exp. 5/22/17
17.2 ml DBCP (W) EPA-1021 lot CH-1351 5000 ug/ml (R4376) exp. 5/31/15
43 ml CS_2 (Acc) APP-9-035-20X lot 212071234 2.0 mg/ml (R4438) exp. 7/18/22
3810 ml MeOH (R4479)
4000 ml DBCP = 43 ug/ml
Other = 21.5 ug/ml LS 2/07/14 Exp. 2/14/14

X1022 Ketones - (Abs) 215 ug/ml

215 ml 82402 lot 070313 2000 ug/ml (R4512) exp. 7/03/16
1785 ml MeOH (R4479)
2000 ml 215 ug/ml LS 2/07/14 Exp. 2/14/14

X1023 ISTD - mix

STM-620-1
Lot: CH-3660
Exp: 12/31/2014
ULTRA 1 mL
Internal Standard Mixture
R4399 3 analyte(s) at 2500 µg/mL in methanol
250 Smith St, No Kingstown, RI 02852 USA
For Laboratory Use Only
LS 2/12/14

80 ml
3920 ml MeOH (R4479)
4000 ml 50 ug/ml LS 2/12/14 Exp. 2/19/14

X1024 Sur - mix

STM-530-1
Lot: CJ-1280
Exp: 05/31/2015
ULTRA 1 mL
Method 8260 Surrogate Standard Mixture
R4397 3 analyte(s) at 2500 µg/mL in methanol
250 Smith St, No Kingstown, RI 02852 USA
For Laboratory Use Only
LS 2/12/14

80 ml
3920 ml MeOH (R4479)
4000 ml 50 ug/ml LS 2/12/14 Exp. 2/19/14

Continued on Page

Read and Understood By

Signed	Date	Signed	Date
LS	2/12/14	TRC	2/13/14

X1025 IC MIX

8260 Mix (U)	MTBE (U)	RMA (CS)	Ketones (RES)	CS2 (ABS)		MEOH	Final Vol.
DWM-589N Lot CK-2561 2000 ug/mL R4502 Exp.09/30/16	STS-440-1 Lot CG-4444 2000 ug/mL R4230 Exp. 01/31/15	SP-1022-345DSZ Lot 2246900 Varied Conc. R4649 Exp. 12/31/14	30006 Lot A090981 5000 ug/mL R4531 Exp. 12/2015	70060 Lot 121813 1000 ug/mL R4661 Exp. 12/18/16		R4479	(uL)
200	200	400	160	400		2640	4000

Ketones = 200 ug/mL
Acrylo/m&p-Xylenes = 200ug/mL
All other = 100ug/ml

Prep: 02/12/14

Exp: 02/19/14

LS 2/12/14

X1026 Gases - Abs

100 ml 30058 lot 010213 2000 ug/ml (R4445) exp. 01/02/16
3900 ml MeOH (R4479)
4000 ml 50 ug/ml LS 2/12/14 Exp. 2/19/14

X1027 ICV MIX

8260 Mix (ABS)	MTBE (S)	RMA (ACC)	Ketones (ABS)	CS2 (ACC)	MEOH	Final Vol.
32001 Lot 022613 2000 ug/mL R4449 Exp.02/26/16	CRM48483 Lot LB94566 2000 ug/mL R4517 Exp. 07/2015	S-2932-R1 Lot 212081711 Varied Conc. R4413 Exp. 9/04/14	82402 Lot 070313 2000 ug/mL R4511 Exp. 07/03/16	APP-9-035-20X Lot 212071234 2.0 mg/mL R4438 Exp. 07/18/22	R4479	(uL)
25	25	250	50	25	1625	2000

Ketones/Acrylo/
m&p-Xylenes/DBCP = 50 ug/mL
OTHER = 25 ug/mL

Prep: 02/13/14

EXP: 02/20/14

LS 2/13/14

X1028 Gases - Ultra

100 ml DWM-544 lot CJ-44469 2000 ug/ml (R4367) exp. 1/31/16
3900 ml MeOH (R4479)
4000 ml 50 ug/ml LS 2/13/14 Exp. 2/20/14

Continued on Page

Read and Understood By

Signed	Date	Signed	Date
LS	2/13/14	TRC	2/17/14

ULTRA SCIENTIFIC
Analytical Solutions

Certificate of Analysis

Rec'd 12/05/13

4-Bromofluorobenzene Solution

Product Number:	STS-110N			**Page:**	1 of 1
Lot Number:	CH-3248	**Lot Issue Date:**	07-Oct-2011	**Expiration Date:**	30-Nov-2015

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
4-bromofluorobenzene	000460-00-4	RM03365	2005 ± 10 µg/mL

Matrix: methanol (methyl alcohol)

Storage: Store at < 4° C

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.

ACCREDITED
ISO 17025:2005 Accredited A2LA Cert. No. 0851-01

ISO 9001:2008 Registered TUV USA, Inc. Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
wwww.ultrasci.com

William J. Leary
Quality Assurance Manager

Certificate of Analysis

Rec'd 1/11/13

ULTRA SCIENTIFIC
Analytical Solutions

Internal Standard Mixture

Product Number: STM-520 **Page:** 1 of 1

Lot Number: CH-3660 **Lot Issue Date:** 15-Nov-2011 **Expiration Date:** 31-Dec-2014

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
chlorobenzene-d5	003114-55-4	RM01430-01	2503 ± 13 µg/mL
1,4-dichlorobenzene-d4	003855-82-1	RM01700-06	2500 ± 13 µg/mL
fluorobenzene	000462-06-6	RM03366-01	2513 ± 13 µg/mL

Matrix: methanol (methyl alcohol)

Storage: Store at < 4° C

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.

A2LA ACCREDITED
ISO 17025:2005 Accredited A2LA Cert. No. 0851-01
ISO 9001:2008 Registered TUV USA, Inc. Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrascl.com

William J. Leary
Quality Assurance Manager

Rec'd 9/18/12

ULTRA SCIENTIFIC
Analytical Solutions

Certificate of Analysis

Method 8260 Surrogate Standard Mixture

Product Number: STM-530 **Page:** 1 of 1

Lot Number: CJ-1280 **Lot Issue Date:** 17-Apr-2012 **Expiration Date:** 31-May-2015

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
4-bromofluorobenzene	000460-00-4	RM03908-	2512 ± 13 µg/mL
dibromofluoromethane	001868-53-7	RM04871-01	2512 ± 13 µg/mL
1,2-dichloroethane-d4	017060-07-0	RM04304-	2512 ± 13 µg/mL
toluene-d8	002037-26-5	RM04138-	2512 ± 13 µg/mL

Matrix: methanol (methyl alcohol)

Storage: Store at < 4° C

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.

A2LA ACCREDITED
ISO 17025:2005 Accredited A2LA Cert. No. 0851-01
ISO 9001:2008 Registered TUV USA, Inc. Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
wwww.ultrascl.com

William J. Leary
Quality Assurance Manager

ULTRA SCIENTIFIC
Analytical Solutions

Certificate of Analysis

Rec'd 8/30/13

VOC Mixture

Product Number: DWM-589N **Page:** 1 of 3

Lot Number: CK-2561 **Lot Issue Date:** 06-Aug-2013 **Expiration Date:** 30-Sep-2016

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
bromochloromethane	000074-97-5	RM00009	2004 ± 10 µg/mL
bromodichloromethane	000075-27-4	RM06861	2008 ± 10 µg/mL
bromoform	000075-25-2	RM00350	2007 ± 10 µg/mL
carbon tetrachloride	000056-23-5	RM00012	2009 ± 10 µg/mL
chloroform	000067-66-3	RM06503	2009 ± 10 µg/mL
dibromochloromethane	000124-48-1	RM04264	2009 ± 10 µg/mL
dibromomethane	000074-95-3	RM02090	2010 ± 10 µg/mL
methylene chloride	000075-09-2	RM06447	1998 ± 10 µg/mL
1,2-dibromoethane	000106-93-4	RM00018	2008 ± 10 µg/mL
1,1-dichloroethane	000075-34-3	RM00019	2006 ± 10 µg/mL
1,2-dichloroethane	000107-06-2	RM04655	2005 ± 10 µg/mL
1,1-dichloroethene	000075-35-4	RM01356	2008 ± 10 µg/mL
cis-1,2-dichloroethene	000156-59-2	RM05327	2005 ± 10 µg/mL
trans-1,2-dichloroethene	000156-60-5	RM00490	2010 ± 10 µg/mL
1,1,1,2-tetrachloroethane	000630-20-6	RM00024	2009 ± 10 µg/mL
1,1,2,2-tetrachloroethane	000079-34-5	RM02540	2009 ± 10 µg/mL
tetrachloroethene	000127-18-4	RM00026	2008 ± 10 µg/mL
1,1,1-trichloroethane	000071-55-6	RM00027	2009 ± 10 µg/mL
1,1,2-trichloroethane	000079-00-5	RM02575	2010 ± 10 µg/mL
trichloroethene	000079-01-6	RM06644	2007 ± 10 µg/mL

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.

ACCREDITED

ISO 17025 Accredited A2LA Cert. No. 0851-01

ISO 9001 Registered TUV USA, Inc. Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 401-295-2330
www.ultrasci.com

William J Leary
Quality Assurance Manager

ULTRA SCIENTIFIC
Analytical Solutions

Certificate of Analysis

VOC Mixture

Product Number: DWM-589N **Page:** 2 of 3

Lot Number: CK-2561 **Lot Issue Date:** 06-Aug-2013 **Expiration Date:** 30-Sep-2016

Analyte	CAS#	Analyte Lot	True Value
1,2-dibromo-3-chloropropane	000096-12-8	RM03703	2006 ± 10 µg/mL
1,2-dichloropropane	000078-87-5	RM02091	2005 ± 10 µg/mL
1,3-dichloropropane	000142-28-9	RM02080	2006 ± 10 µg/mL
2,2-dichloropropane	000594-20-7	RM04312	2005 ± 10 µg/mL
1,1-dichloropropene	000563-58-6	RM01234	2003 ± 10 µg/mL
cis-1,3-dichloropropene	010061-01-5	RM01442	2003 ± 10 µg/mL
trans-1,3-dichloropropene	010061-02-6	RM01443	2005 ± 10 µg/mL
hexachlorobutadiene	000087-68-3	RM00438	2006 ± 10 µg/mL
1,2,3-trichloropropane	000096-18-4	RM02562	2008 ± 10 µg/mL
benzene	000071-43-2	RM03830	2005 ± 10 µg/mL
n-butylbenzene	000104-51-8	RM03651	2000 ± 10 µg/mL
sec-butylbenzene	000135-98-8	RM01800	2003 ± 10 µg/mL
tert-butylbenzene	000098-06-6	RM00043	2009 ± 10 µg/mL
ethylbenzene	000100-41-4	RM00783	2004 ± 10 µg/mL
isopropylbenzene	000098-82-8	RM00835	2009 ± 10 µg/mL
4-isopropyltoluene	000099-87-6	RM00046	2001 ± 10 µg/mL
naphthalene	000091-20-3	RM02406	2005 ± 10 µg/mL
n-propylbenzene	000103-65-1	RM02468	2008 ± 10 µg/mL
styrene	000100-42-5	RM04479	2002 ± 10 µg/mL
toluene	000108-88-3	RM02555	2009 ± 10 µg/mL
1,2,4-trimethylbenzene	000095-63-6	RM00050	2004 ± 10 µg/mL
1,3,5-trimethylbenzene	000108-67-8	RM02565	2010 ± 10 µg/mL
o-xylene	000095-47-6	RM00052	2002 ± 10 µg/mL
m-xylene	000108-38-3	RM00053	2002 ± 10 µg/mL

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.

ACCREDITED

ISO 17025 Accredited A2LA Cert. No. 0851-01

ISO 9001 Registered TUV USA, Inc. Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 401-295-2330
www.ultrasci.com

William J. Leary
Quality Assurance Manager

ULTRA SCIENTIFIC
Analytical Solutions

Certificate of Analysis

VOC Mixture

Product Number: DWM-589N **Page:** 3 of 3

Lot Number: CK-2561 **Lot Issue Date:** 06-Aug-2013 **Expiration Date:** 30-Sep-2016

Analyte	CAS#	Analyte Lot	True Value
p-xylene	000106-42-3	RM02647	2006 ± 10 µg/mL
bromobenzene	000108-86-1	RM00056	2008 ± 10 µg/mL
chlorobenzene	000108-90-7	RM01874	2008 ± 10 µg/mL
2-chlorotoluene	000095-49-8	RM03906	2010 ± 10 µg/mL
4-chlorotoluene	000106-43-4	RM01866	2009 ± 10 µg/mL
1,2-dichlorobenzene	000095-50-1	NT00355	2004 ± 10 µg/mL
1,3-dichlorobenzene	000541-73-1	NT00356	2003 ± 10 µg/mL
1,4-dichlorobenzene	000106-46-7	RM04303	2008 ± 10 µg/mL
1,2,3-trichlorobenzene	000087-61-6	RM02563	2009 ± 10 µg/mL
1,2,4-trichlorobenzene	000120-82-1	RM00063	2008 ± 10 µg/mL

Matrix: methanol (methyl alcohol)

Storage: Store at < 4° C

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.

ACCREDITED

ISO 17025 Accredited A2LA Cert. No. 0851-01

ISO 9001 Registered TUV USA, Inc. Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 401-295-2330
www.ultrasci.com

William J. Leary
Quality Assurance Manager

Rec'd 9/18/12

ULTRA SCIENTIFIC
Analytical Solutions

Certificate of Analysis

Methyl tert-Butyl Ether Solution

Product Number: STS-440 **Page:** 1 of 1

Lot Number: CG-4444 **Lot Issue Date:** 03-Jan-2011 **Expiration Date:** 31-Jan-2015

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
tert-butylmethyl ether	001634-04-4	46160	2007 ± 10 µg/mL

Matrix: methanol (methyl alcohol)

Storage: Store at < 4° C

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.

ACCREDITED

ISO 17025:2005 Accredited A2LA Cert. No. 0851-01

ISO 9001:2008 Registered TUV USA, Inc. Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com

William J. Leary
Quality Assurance Manager

Rec'd 12/16/13

660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com • www.chemservice.com

CERTIFICATE OF ANALYSIS

Mixture of 3 analytes

CONCENTRATION	Varied concentration in Methanol
CATALOG NUMBER	SP-1022345DSZ-1ML
LOT NUMBER	2246900
DATE CERTIFIED	12/10/13
EXPIRATION DATE	12/31/14
STORAGE	Room Temperature
HANDLING	See MSDS
INTENDED USE	For Laboratory Use Only

ID	Analyte	CAS	Weight Analyte (mg)	Lot	Purity	Certified Concentration (ug/mL)
N-11686	Dicyclopentadiene	77-73-6	25.600	00011137	99.00	1013.8
N-15088	Bicyclo[2,2,1]hepta-2,5-diene	121-46-0	25.400	396800	99.50	1010.9
N-11034	Acrylonitrile	107-13-1	51.100	00008601	99.50	2033.8

Analytical Test	Value
CONCENTRATION (GC/FID)	VERIFIED

Instructions for Use:
Shake mixture prior to use. If particles are present, sonicate for homogeneity. If sample is diluted to lower concentrations, Class A volumetric glassware must be used.
Minimum Sample Size- 0.2 uL for Direct Injection.
Chem Service Inc. guarantees the expanded uncertainty of the above analytes to be +/- 2.0% of the certified concentrations based on gravimetric preparation. The test results published in this report were obtained using equipment capable of producing results that are traceable to NIST and through NIST to the International System of Units (SI). The reported expanded uncertainty of measurement is stated as the combined standard uncertainty of measurement multiplied by the coverage factor k (k=2) such that the coverage probability corresponds to approximately 95%. For certified reference materials, homogeneity and thermal stability testing are available upon request.

Certified By:

Mary Beth O'Donnell

Mary Beth O'Donnell
CSM/TC

COA Form
Revision 1 (01/2013)

ISO 9001
Certificate Number:31610

ISO/IEC 17025
Accreditation Number:63520

Reference Material Producer
ISO GUIDE 34
Accreditation Number:63520

660 Tower Lane • P.O. Box 599 • West Chester, PA 19381-0599
1-800-452-9994 • 1-610-692-3026 • Fax 1-610-692-8729
info@chemservice.com * www.chemservice.com

CERTIFICATE OF ANALYSIS

Gas Chromatography / Flame Ionization Detector (GC/FID)

Data file:	C:\CHEM32\1\DATA\1213\5\G1004643.D		
Sample name:	SP-102245052		
Instrument:	GC 1	Sample type:	Sample
Injection date:	12/10/2013 9:02:35 AM	Location:	Vial 1
Acq. method:	MIX1.M	Injection volume:	1.0uL
Column name:	DB-624 (30m x 0.53mm x 3.0um)		



Signal: FID1 A, Front Signal

RT [min]	Type	Width [min]	Area	Height	Area%
3.573	BB	0.0369	1294.2858	543.3721	37.2976
5.982	BB	0.0547	1025.1239	303.9188	29.5412
11.980	BB	0.0531	1150.7445	334.8539	33.1612
		Sum	3470.1542		

ISO 9001 Certificate Number:31610 | ISO/IEC 17025 Accreditation Number:63520 | ISO GUIDE 34 Accreditation Number:63520

RESTEK

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Rec'd 8/29/13

ACCREDITED
ISO Guide 34 Accredited Reference Material Producer Certificate #3222.01

ilac-MRA

ACCREDITED
ISO/IEC 17025 Accredited Testing Laboratory Certificate #3222.02

Certificate of Analysis

FOR LABORATORY USE ONLY-READ MSDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 30006 **Lot No.:** A090981

Description : VOA Calibration Mix #1

Calibration Std #1 ketones 5000µg/mL, P&T Methanol/Water(90:10), 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 2015 **Storage:** 0°C or colder

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)		
1	Acetone	5,000.0 µg/mL	+/- 29.0704	µg/mL	Gravimetric
	CAS # 67-64-1		+/- 399.3449	µg/mL	Unstressed
	Purity 99%		+/- 399.5405	µg/mL	Stressed
2	2-Butanone (MEK)	5,000.0 µg/mL	+/- 29.0704	µg/mL	Gravimetric
	CAS # 78-93-3		+/- 399.3449	µg/mL	Unstressed
	Purity 99%		+/- 399.5405	µg/mL	Stressed
3	4-Methyl-2-pentanone (MIBK)	5,000.0 µg/mL	+/- 29.0704	µg/mL	Gravimetric
	CAS # 108-10-1		+/- 399.3449	µg/mL	Unstressed
	Purity 99%		+/- 399.5405	µg/mL	Stressed
4	2-Hexanone	5,000.0 µg/mL	+/- 29.0704	µg/mL	Gravimetric
	CAS # 591-78-6		+/- 399.3449	µg/mL	Unstressed
	Purity ----%		+/- 399.5405	µg/mL	Stressed

Solvent: P&T Methanol/Water (90:10)
CAS # 67-56-1/7732-18-5
Purity 99%

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Rec'd 2/19/14

ISO 9001 QS Registered
ISO 17025-34-35-43 Accredited
Scopes: http://AbsoluteStandards.com

CERTIFIED WEIGHT REPORT

			Lot #	
Part Number:	70060	**Solvent(s):**	DG596	Methanol
Lot Number:	121813			
Description:	Carbon disulphide			
Expiration Date:	121816			
Recommended Storage:	Refrigerate (4 °C)			
Nominal Concentration (µg/mL):	1000			

Formulated By:	Paul Barron	121813 DATE
Reviewed By:	Pedro L. Rentas	121813 DATE

5E-05 Balance Uncertainty

Weight(s) shown below were combined and diluted to: 200.0 0.015 Flask Uncertainty

MSDS Information
(Solvent Safety Info. On Attached pg.)

	Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty	CAS#	OSHA PEL (TWA)	LD50
1.	Carbon disulphide	60	PO 12273PO	1000	99	0.2	0.20203	0.20225	1001.1	0.0041	00075-15-0	4 ppm (12mg/m3) (skin)	orl-rat 1200mg/kg

Method GC6MSD-1: Column: (Vocol 60m X 0.25mm ID X 1.5µm film thickness). Temp. 1 = 35°C (10min.), Temp. 2 = 200°C (8.75 min.), Rate = 4°C/min., Injector Temp.= 200°C, Detector Temp. = 220°C. Analysis performed by Candice Warren.





Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ISO 9001 QS Registered
ISO 17025-34-35-43 Accredited
Scopes: http://AbsoluteStandards.com

Run 34, "P70060 L121813 [1000µg/mL in MeOH]"

Run Length: 59.99 min, 35997 points at 10 points/second.
Created: Sat, Dec 28, 2013 at 1:37:07 AM.
Sampled: Sequence "122613-GC1", Method "GC1-M7".
Analyzed using Method "GC1-M7".

Analyte	PID RT - (min.)
Carbon disulphide	10.81

Comments
GC1-M7 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow = 150mL/min., Helium (carrier) = 10mL/min.,
Helium (make-up) = 40mL/min., Hydogen (make-up) = 100mL/min.,
Oven Profile: Temp. 1 = 35°C (Time 1 = 10 min.), Temp 2 = 200°C (Time 2 = 8.75 min.),
Rate = 4°C/min., Total run time = 60 min. Injector temp. = 200°C, PID Temp. = 200°C.
ELCD Signal = Edaq Channel 1 PID Signal = Edaq Channel 2
Standard injection = 0.5µL, Range=4 Purge Valve = 8 min



Absolute Sta ırds, Inc.
800-368-1131
www.absolutestandards.com

Certified Re ence Material CRM*

Recd 1/14/13

ISO 90 S Registered
ISO 17025-34 43 Accredited
Scopes: http://AbsoluteStandards.com

CERTIFIED WEIGHT REPORT

Part Number:	30058	**Solvent(s):**	Lot # DE997	Methanol
Lot Number:	010213			
Description:	EPA Method 502/524 - Volatile Gases Mix #1			
	6 components			
Expiration Date:	010216			
Recommended Storage:	Freezer (0 °C)			
Nominal Concentration (µg/mL):	2000			

Weight(s) shown below were combined and diluted to: 499.9 (5E-05 Balance Uncertainty) 0.058 Flask Uncertainty

Formulated By:	Vincent K. Criscio, Jr.	010213 DATE
Reviewed By:	Pedro L. Rentas	010213 DATE

MSDS Information
(Solvent Safety Info. On Attached pg.)

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Uncertainty Purity	Target Weight (g)	Actual Weight (g)	Actual Conc(µg/mL)	Expanded Uncertainty	CAS#	OSHA PEL (TWA)	LD50
1. Bromomethane	50	01611JX	2000	99.5	0.2	1.00475	1.0069	2004.3	0.00403	00074-83-9	5 ppm (20mg/m3/8H) (skin)	orl-rat 214mg/kg
2. Chloroethane	72	00727KX	2000	99	0.2	1.00982	1.0100	2000.4	0.00405	00075-00-3	1000 ppm (2600mg/m3/8H)	N/A
3. Chloromethane	79	06908MS	2000	99.5	0.2	1.00475	1.0117	2013.8	0.00403	00074-87-3	100 ppm	orl-rat 1800mg/kg
4. Dichlorodifluoromethane	134	92-0487	2000	99	0.2	1.00982	1.0204	2020.9	0.00405	00075-71-8	1000 ppm (4950mg/m3/8H)	N/A
5. Trichlorofluoromethane	294	01823MW	2000	99	0.2	1.00982	1.0150	2010.3	0.00405	00075-69-4	1000 ppm (5600mg/m3/8H)	ipr-mus 1743mg/kg
6. Vinyl chloride	305	1281633 21306138	2000	99.5	0.2	1.00475	1.0113	2013.0	0.00403	00075-01-4	N/A	N/A

Run 9, "P30058 L010213 [2000µg/mL in MeOH]"

Run Length: 15.00 min. 8999 points at 10 points/second.
Created: Wed, Jan 2, 2013 at 5:05:03 PM.
Sampled: Sequence "123112-GC1", Method "GC1-M9".
Analyzed using Method "GC1-M9".

Comments
GC1-M9 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow = 150mL/min., Helium (carrier) = 10mL/min.,
Helium (make-up) = 40mL/min., Hydogen (make-up) = 100mL/min.,
Oven Profile: Temp. 1 = 35°C (Time 1 = 9 min.), Temp 2 = 200°C (Time 2 = 1 min.),
Rate = 33°C/min., Total run time = 15 min. Injector temp. = 200°C, FID Temp. = 200°C.
ELCD Signal = Edaq Channel 2 PID Signal = Edaq Channel 2
Standard injection = 0.5µL, Range=4 Purge Valve = 0 min



Analyte	ELCD RT (min.)
Dichlorodifluoromethane *	5.34
Chloromethane	6.03
Vinyl chloride	6.34
Bromomethane	7.67
Chloroethane	7.93
Trichlorofluoromethane	8.76

* +/- 20%CRM

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Rec'd 3/19/13

ISO 9001 QS Registered
ISO 17025-34-35-43 Accredited
Scopes: http://AbsoluteStandards.com

CERTIFIED WEIGHT REPORT

		Lot #	Solvent(s):
Part Number:	32001	DE997Q2	Methanol
Lot Number:	022613		
Description:	EPA Method 502.2/524.2 Volatile Organic Compounds 54 components		
Expiration Date:	022616		
Recommended Storage:	Freezer (0 °C)		
Nominal Concentration (µg/mL):	2000		

Formulated By:	Justin Dippold	022613	DATE
Reviewed By:	Pedro L. Rentas	022613	DATE

5E-05 Balance Uncertainty

Volume(s) shown below were combined and diluted to: 200.0 0.015 Flask Uncertainty

Compound	Part Number	Lot Number	Dil. Factor	Initial Vol. (mL)	Uncertainty Pipette	Initial Conc.(ug/mL)	Final Conc.(ug/mL)	Expanded Uncertainty	CAS#	MSDS Information (Solvent Safety Info. On Attached pg.) OSHA PEL (TWA)	LD50
1. Bromodichloromethane	30171	122112	0.10	20.00	0.011	20016.1	2001.5	0.00408	00075-27-4	N/A	orl-rat 916mg/kg
2. Dibromochloromethane	30171	122112	0.10	20.00	0.011	20011.7	2001.1	0.00408	00124-48-1	N/A	orl-rat 848mg/kg
3. cis-1,2-Dichloroethene	30171	122112	0.10	20.00	0.011	20018.4	2001.7	0.00404	00156-59-2	N/A	N/A
4. trans-1,2-Dichloroethene	30171	122112	0.10	20.00	0.011	20001.9	2000.1	0.00408	00156-60-5	N/A	orl-rat 1235mg/kg
5. Methylene chloride	30171	122112	0.10	20.00	0.011	20004.5	2000.4	0.00401	00075-09-2	500 ppm	orl-rat 2136mg/kg
6. 1,1-Dichloroethene	32251	122112	0.10	20.00	0.011	20002.4	2000.1	0.00404	00075-35-4	1 ppm (4mg/m3/8H)	orl-rat 200mg/kg
7. Bromochloromethane	30160	102011	0.10	20.00	0.011	20007.7	2000.7	0.00404	00074-97-5	200 ppm (1050mg/m3/8H)	orl-rat 5000mg/kg
8. Bromoform	30160	102011	0.10	20.00	0.011	20007.6	2000.7	0.00404	00075-25-2	0.5 ppm (5mg/m3) (skin)	orl-rat 1147mg/kg
9. Carbon tetrachloride	30160	102011	0.10	20.00	0.011	20008.4	2000.8	0.00402	00056-23-5	2 ppm (12.6mg/m3/8H)	orl-rat 2350mg/kg
10. Chloroform	30160	102011	0.10	20.00	0.011	20006.5	2000.6	0.00400	00067-66-3	50 ppm (240mg/m3) (CL)	orl-rat 908mg/kg
11. Dibromomethane	30160	102011	0.10	20.00	0.011	20007.7	2000.7	0.00404	00074-95-3	N/A	orl-rat 108mg/kg
12. 1,1-Dichloroethane	30160	102011	0.10	20.00	0.011	20005.3	2000.4	0.00404	00075-34-3	100 ppm	orl-rat 725mg/kg
13. 2,2-Dichloropropane	30160	102011	0.10	20.00	0.011	20011.0	2001.0	0.00408	00594-20-7	N/A	N/A
14. Tetrachloroethene	30160	102011	0.10	20.00	0.011	20017.2	2001.6	0.00404	00127-18-4	5 ppm (170mg/m3/8H)(fina	orl-rat 2629mg/kg
15. 1,1,1-Trichloroethane	30160	102011	0.10	20.00	0.011	20014.4	2001.3	0.00404	00071-55-6	350 ppm (1900mg/m3/8H)	orl-rat 10300mg/kg
16. 1,2-Dibromo-3-chloropropane	30161	072312	0.10	20.00	0.011	20020.9	2002.0	0.00404	00096-12-8	0.001 ppm	orl-rat 170mg/kg
17. 1,2-Dibromoethane	30161	072312	0.10	20.00	0.011	20023.4	2002.2	0.00404	00106-93-4	20 ppm (8H)	orl-rat 108mg/kg
18. 1,2-Dichloroethane	30161	072312	0.10	20.00	0.011	20004.5	2000.4	0.00401	00107-06-2	50 ppm (8H)	orl-rat 670mg/kg
19. 1,2-Dichloropropane	30161	072312	0.10	20.00	0.011	20022.9	2002.2	0.00404	00078-87-5	75 ppm (350mg/m3/8H)	orl-rat 1947mg/kg
20. 1,3-Dichloropropane	30161	072312	0.10	20.00	0.011	20007.0	2000.6	0.00404	00142-28-9	N/A	unr-mus 3600mg/kg
21. 2-Dichloropropene	30161	072312	0.10	20.00	0.011	20022.3	2002.1	0.00404	00563-58-6	N/A	N/A
22. cis-1,3-Dichloropropene	30161	072312	0.10	20.00	0.011	20014.4	2001.3	0.00404	10061-01-5	N/A	N/A
23. trans-1,3-Dichloropropene	30161	072312	0.10	20.00	0.011	20012.2	2001.1	0.00404	10061-02-6	N/A	N/A
24. Hexachloro-1,3-butadiene	30161	072312	0.10	20.00	0.011	20007.3	2000.6	0.00408	00087-68-3	0.02 ppm (0.24mg/m3/8H)	orl-rat 90mg/kg
25. 1,1,1,2-Tetrachloroethane	30161	072312	0.10	20.00	0.011	20007.5	2000.7	0.00404	00630-20-6	N/A	orl-rat 670mg/kg
26. 1,1,2,2-Tetrachloroethane	30161	072312	0.10	20.00	0.011	20008.0	2000.7	0.00404	00079-34-5	5 ppm (35mg/m3/9H)(skin)	orl-rat 800mg/kg
27. 1,1,2-Trichloroethane	30161	072312	0.10	20.00	0.011	20008.2	2000.7	0.00408	00079-00-5	10 ppm (45mg/m3/8H)(skin	orl-rat 836mg/kg
28. Trichloroethene	30161	072312	0.10	20.00	0.011	20012.5	2001.2	0.00402	00079-01-6	50 ppm (270mg/m3/8H)	orl-mus 2402mg/kg
29. 1,2,3-Trichloropropane	30161	072312	0.10	20.00	0.011	20009.5	2000.9	0.00404	00096-18-4	10 ppm (60mg/m3/8H)	orl-rat 320mg/kg
30. Benzene	30162	102711	0.10	20.00	0.011	20014.6	2001.4	0.00415	00071-43-2	1 ppm	orl-rat 4894mg/kg
31. Bromobenzene	30162	102711	0.10	20.00	0.011	20033.7	2003.3	0.00418	00108-86-1	N/A	orl-rat 2699mg/kg
32. n-Butyl benzene	30162	102711	0.10	20.00	0.011	20007.5	2000.7	0.00418	00104-51-8	N/A	N/A
33. Ethyl benzene	30162	102711	0.10	20.00	0.011	20004.9	2000.4	0.00415	00100-41-4	100 ppm (435mg/m3/8H)	orl-rat >2000mg/kg
34. p-Isopropyl toluene	30162	102711	0.10	20.00	0.011	20026.0	2002.5	0.00418	00099-87-6	N/A	orl-rat 4750mg/kg
35. Naphthalene	30162	102711	0.10	20.00	0.011	20011.8	2001.1	0.00418	00091-20-3	10 ppm (50mg/m3/8H)	orl-rat 490mg/kg
36. Styrene	30162	102711	0.10	20.00	0.011	20011.3	2001.0	0.00418	00100-42-5	100 ppm	orl-rat 5000mg/kg
37. Toluene	30162	102711	0.10	20.00	0.011	20007.1	2000.6	0.00418	00108-88-3	200 ppm	orl-rat 5000mg/kg
38. 1,2,3-Trichlorobenzene	30162	102711	0.10	20.00	0.011	20028.3	2002.7	0.00418	00087-61-6	N/A	ipr-mus 1390mg/kg
39. 1,2,4-Trichlorobenzene	30162	102711	0.10	20.00	0.011	20016.6	2001.6	0.00418	00120-82-1	5 ppm (CL) (40mg/m3)	orl-rat 756mg/kg
40. 1,2,4-Trimethylbenzene	30162	102711	0.10	20.00	0.011	20014.7	2001.4	0.00422	00095-63-6	N/A	orl-rat 5g/kg
41. 1,3,5-Trimethylbenzene	30162	102711	0.10	20.00	0.011	20011.3	2001.0	0.00422	00108-67-8	N/A	N/A
42. m-Xylene	30162	102711	0.10	20.00	0.011	20013.4	2001.3	0.00418	00108-38-3	100 ppm (435mg/m3/8H)	orl-rat 5g/kg
43. tert-Butyl benzene	30163	102011	0.10	20.00	0.011	20002.3	2000.1	0.00404	00098-06-6	N/A	N/A
44. sec-Butyl benzene	30163	102011	0.10	20.00	0.011	20001.8	2000.1	0.00404	00135-98-8	N/A	orl-rat 2240mg/kg
45. Chlorobenzene	30163	102011	0.10	20.00	0.011	20006.7	2000.6	0.00404	00108-90-7	75 ppm (350mg/m3/8H)	orl-rat 2290mg/kg
46. 2-Chlorotoluene	30163	102011	0.10	20.00	0.011	20006.4	2000.5	0.00404	00095-49-8	50 ppm (250mg/m3/8H)	unr-rat 5700mg/kg
47. 4-Chlorotoluene	30163	102011	0.10	20.00	0.011	20003.7	2000.3	0.00404	00106-43-4	N/A	orl-rat 2100mg/kg
48. 1,2-Dichlorobenzene	30163	102011	0.10	20.00	0.011	20003.1	2000.2	0.00404	00095-50-1	50 ppm (300mg/m3) (CL)	orl-rat 500mg/kg
49. 1,3-Dichlorobenzene	30163	102011	0.10	20.00	0.011	20003.8	2000.3	0.00408	00541-73-1	N/A	ipr-mus 1062mg/kg
50. 1,4-Dichlorobenzene	30163	102011	0.10	20.00	0.011	20029.8	2002.9	0.00404	00106-46-7	75 ppm (450mg/m3/8H)	orl-rat 500mg/kg
51. Isopropylbenzene	30163	102011	0.10	20.00	0.011	20006.1	2000.5	0.00408	00098-82-8	50 ppm (245mg/m3/8H)	orl-rat 1400mg/kg
52. n-Propylbenzene	30163	102011	0.10	20.00	0.011	20013.7	2001.3	0.00408	00103-65-1	N/A	orl-rat 6040mg/kg
53. o-Xylene	30163	102011	0.10	20.00	0.011	20008.2	2000.7	0.00408	00095-47-6	100 ppm (435mg/m3/8H)	ipr-mus 1364mg/kg
54. p-Xylene	30163	102011	0.10	20.00	0.011	20007.6	2000.7	0.00404	00106-42-3	100 ppm (435mg/m3/8H)	orl-rat 5g/kg

Run 37, "P32001 L022613 [2000µg/mL in MeOH]-PID"

Run Length: 59.99 min, 35997 points at 10 points/second.
Created: Tue, Feb 26, 2013 at 6:36:10 PM.
Sampled: Sequence "022313-GC1", Method "GC1-M7".
Analyzed using Method "GC1-M7".

Comments
GC1-M7 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow = 150mL/min., Helium (carrier) = 10mL/min.,
Helium (make-up) = 40mL/min., Hydogen (make-up) = 100mL/min.,
Oven Profile: Temp. 1 = 35°C (Time 1 = 10 min.), Temp 2 = 200°C (Time 2 = 8.75 min.),
Rate = 4°C/min., Total run time = 60 min. Injector temp. = 200°C, PID Temp. = 200°C.
ELCD Signal = Edaq Channel 1 PID Signal = Edaq Channel 2
Standard Injection = 0.5µL, Range=4 Purge Valve = 8 min



Analyte	PID RT (min.)	ELCD RT (min.)
1,1-Dichloroethene	11.23	11.25
Methylene chloride	NA	13.28
trans-1,2-Dichloroethene	14.32	14.34
1,1-Dichloroethane	NA	16.06
2,2-Dichloropropane	NA	18.11
cis-1,2-Dichloroethene	18.33	18.35
Chloroform	NA	18.92
Bromochloromethane	NA	19.55
1,1,1-Trichloroethane	NA	20.29
1,1-Dichloropropene	20.81	20.84
Carbon tetrachloride	NA	21.16
1,2-Dichloroethane	NA	21.88
Benzene	21.81	NA
Trichloroethene	23.92	23.94
1,2-Dichloropropane	NA	24.62
Bromodichloromethane	NA	25.48
Dibromomethane	NA	25.68
cis-1,3-Dichloropropene	27.34	27.36
Toluene	28.38	NA
trans-1,3-Dichloropropene	29.11	29.14
1,1,2-Trichloroethane	NA	29.67
1,3-Dichloropropane	NA	30.59
Tetrachloroethene	30.59	30.59
Dibromochloromethane	NA	31.47
1,2-Dibromoethane	NA	32.15
Chlorobenzene	33.55	33.57
1,1,1,2-Tetrachloroethane	NA	33.73
Ethylbenzene	33.70	NA
m-Xylene	33.96	NA
p-Xylene	33.96	NA
o-Xylene	35.51	NA
Styrene	35.68	NA
Isopropylbenzene *	36.78	NA
Bromoform	NA	36.98
1,1,2,2-Tetrachloroethane	NA	37.54
1,2,3-Trichloropropane	NA	38.06
n-Propylbenzene	38.22	NA
Bromobenzene	38.42	38.45
1,3,5-Trimethylbenzene	38.78	NA
2-Chlorotoluene	38.87	38.90
4-Chlorotoluene	39.05	39.07
tert-Butylbenzene	40.06	NA
1,2,4-Trimethylbenzene	40.20	NA
sec-Butylbenzene	40.82	NA
p-Isopropyltoluene	41.33	NA
1,3-Dichlorobenzene	41.71	41.73
1,4-Dichlorobenzene	42.11	42.13
n-Butylbenzene	42.83	NA
1,2-Dichlorobenzene	43.47	43.49
1,2-Dibromo-3-chloropropane	NA	46.44
1,2,4-Trichlorobenzene	49.58	49.60
Hexachlorobutadiene	50.06	50.09
Naphthalene	50.58	NA
1,2,3-Trichlorobenzene	51.49	51.52

Run 37, "P32001 L022613 [2000µg/mL in MeOH]-ELCD"

Run Length: 59.99 min, 35997 points at 10 points/second.
Created: Tue, Feb 26, 2013 at 6:36:10 PM.
Sampled: Sequence "022313-GC1", Method "GC1-M7".
Analyzed using Method "GC1-M7".

Comments
GC1-M7 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow = 150mL/min., Helium (carrier) = 10mL/min.,
Helium (make-up) = 40mL/min., Hydogen (make-up) = 100mL/min.,
Oven Profile: Temp. 1 = 35°C (Time 1 = 10 min.), Temp 2 = 200°C (Time 2 = 8.75 min.),
Rate = 4°C/min., Total run time = 60 min. Injector temp. = 200°C, PID Temp. = 200°C.
ELCD Signal = Edaq Channel 1 PID Signal = Edaq Channel 2
Standard Injection = 0.5µL, Range=4 Purge Valve = 8 min



Recd 9/28/13

TraceCERT®
Traceable Certified Reference Materials

SUPELCO Analytical
595 North Harrison Road · Bellefonte, PA
16823-0048 USA · Phone(814)359-3441

Certificate of Analysis

Description: Methyl tert-butyl ether, 1x1ml, 2000µg/ml in methanol

Catalog Number: CRM48483

Lot Number: LB94566

Expiration Date: July 2015

Storage: Room Temperature

Analytical Method Parameters:
Column: VOCOL 105m x 0.53mm x 3.0µm df
50°C for 4 min to 160°C at 10°C/min
Detector: FID, 310°C
Injection Volume: 1.0uL



Elution	Analyte	Lot Number	CAS Number	Chromatographic Purity % (1)	Certified Gravimetric Conc. µg/ml (2)	Expanded Uncertainty µg/ml (3)	Analytical Conc. µg/ml (2)
1.	Methyl tert-Butyl ether	LB94137	1634-04-4	99.9	1999	5.89	2061
	Internal Standard	N/A	N/A	N/A	N/A	N/A	N/A

ACLASS ACCREDITED TESTING Cert# AT-1607

ACLASS ACCREDITED REFERENCE MATERIAL PRODUCER Cert# AR-1606

Produced in double accredited laboratory fulfulling
ISO/IEC 17025 and
ISO Guide 34

SIGMA-ALDRICH®

Notes:

• 1. All starting materials are Certified Reference Materials.

• 2. NIST traceable weights are used to verify balance calibration with the preparation of each lot. Concentration of analyte in solution is ug/ml. Weights are corrected for analytes less than 98% pure.

• 3. Certified value is the gravimetric concentration weighted at room temperature. Uncertainty of the gravimetric concentration in this document are expressed as Expanded Uncertainty (Ucrm) corresponding to the 95% confidence interval. Ucrm is derived from the combined standard uncertainty multiplied by the coverage factor k = 2. The components of combined standard uncertainty include the uncertainties due to characterization, production, homogeneity, long-term stability and short-term stability. The components due to stability are generally considered to be negligible unless otherwise indicated by stability studies.

• Homogeneity was assessed in accordance with ISO Guide 35. Completed units were sampled using a random stratified sampling protocol. The results of chemical analysis were then compared by Single Factor Analysis of Variance (ANOVA). The uncertainty due to homogeneity was derived from the ANOVA. Heterogeneity was not detected under the conditions of the ANOVA.

• Product intended for laboratory use only. Supelco warrants that its products conform to the information contained in this publication. Purchaser must determine the suitability of the product for its particular use. Please see the latest catalog or order invoice and packing slip for additional terms and conditions of sale.



Test Date: August 02, 2012
Form: CRM48483

Duane Funk

Duane Funk
QC Manager

ACLASS ACCREDITED TESTING Cert# AT-1607

ACLASS ACCREDITED REFERENCE MATERIAL PRODUCER Cert# AR-1606

Produced in double accredited laboratory fulfilling
ISO/IEC 17025 and
ISO Guide 34

SIGMA-ALDRICH®

125 Market Street
New Haven, CT 06513
USA

AccuStandard® Inc.

Rec'd 2/20/14

Tel (203)786-5290
Fax (203)786-5287
www.AccuStandard.com

CERTIFICATE OF ANALYSIS

CATALOG NO: S-2932-R1
DESCRIPTION: RMA Compounds
LOT: 212081711
SOLVENT: Methanol

EXPIRATION: Sep 4,2014
DATE CERTIFIED: Sep 4, 2012
SAMPLE SIZE: 1.5 mL
STORAGE CONDITION: Refrig (0-5° C)
HAZARDS: HIGHLY FLAMMABLE

Refer to the MSDS for additional safety information

[X] Included on ISO/IEC 17025 Scope of Accreditation
[X] Included on ISO Guide 34 Scope of Accreditation

Component	Cas Number	Purity % (GC/FID)	Prepared Concentration[1] (µg/mL)	Certified Analyte Concentration[2] (µg/mL)
Acrylonitrile	107-13-1	99.8	400.4	399.6
Bicyclo[2,2,1]hepta-2,5-diene	121-46-0	99.1	201.2	199.4
1,2-Dibromo-3-chloropropane	96-12-8	100	200.7	200.7
Dicyclopentadiene	77-73-6	98.9	201.6	199.4

4 Components

1. All weights are traceable through NIST, Test No. 822-275872-11
2. Certified Analyte Concentration = Purity x Prepared Concentration. The Uncertainty associated with the gravimetric values reported on this certificate is ±0.24%. The CRM Uncertainty calculated for this product is ±5%. These values are the expanded uncertainty and represent an estimated standard deviation equal to the positive square root of the total variation of the uncertainty of components. A normal distribution is assumed and a coverage factor of K=2 is chosen using approximately a 95% confidence level.
3. A product with a suffix (-1A, -2B, etc. or -01, -02, etc.) on its lot# has had its expiration date extended and is identical to the same lot# without the suffix.

For use in routine laboratory analysis.

Certified by: R. Cooper
Russ Cooper, QC Manager

AccuStandard is accredited to ISO Guide 34, ISO/IEC 17025 and certified to ISO 9001

OR-ORG/INO-001
Rev. 7/11

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

Rec'd 8/30/13

ISO 9001:2008 Registered
ISO 17025-34 ...-43 Accredited
Scopes: http://AbsoluteStandards.com

CERTIFIED WEIGHT REPORT

		Lot #	Solvent(s):	
Part Number:	82402	DH247	Methanol	(90%)
Lot Number:	070313	053113	Water	(10%)
Description:	EPA Method 8240A - Mix #2			
	Ketones 4 components			
Expiration Date:	070316			
Recommended Storage:	Refrigerate (4 °C)			
Nominal Concentration (µg/mL):	2000			

5E-05 Balance Uncertainty

Volume(s) shown below were combined and diluted to: 200.0 0.015 Flask Uncertainty

Formulated By:	Paul Barron	070313 DATE
Reviewed By:	Pedro L. Rentas	070313 DATE

MSDS Information
(Solvent Safety Info. On Attached pg.)

Compound	Part Number	Lot Number	Dil. Factor	Initial Vol. (mL)	Uncertainty Pipette	Initial Conc.(ug/mL)	Final Conc.(ug/mL)	Expanded Uncertainty	CAS#	OSHA PEL (TWA)	LD50
1. Acetone	82492	100412	0.10	20.00	0.011	20009.9	2000.9	0.00402	00067-64-1	750 ppm	orl-rat 9750mg/kg
2. 2-Butanone (Methyl ethyl ketone)	82492	100412	0.10	20.00	0.011	20011.9	2001.1	0.00400	00078-93-3	200 ppm (590mg/m3/8H)	orl-rat 2737mg/kg
3. 2-Hexanone	82492	100412	0.10	20.00	0.011	20010.0	2000.9	0.00408	00591-78-6	100 ppm (410mg/m3/8H)	orl-rat 2590mg/kg
4. 4-Methyl-2-pentanone	82492	100412	0.10	20.00	0.011	20001.5	2000.1	0.00402	00108-10-1	100 ppm (410mg/m3/8H)	orl-rat 2080mg/kg



Method: GC6MSD-1. Detector: Mass Selective Detector. Column: Vocol (60m X 0.25mm ID X 1.5µm film thickness). Oven Profile: Temp. 1 = 35°C (Time 1 = 10min.), Temp. 2=200°C (Time 2 = 8.75 min.), Rate = 4°C/min., Injector Temp. = 200°C, Detector Temp. = 220°C. Analyst: Candice Warren.

Name	MSD RT (min.)
Acetone	10.61
2-Butanone	18.17
4-Methyl-2-pentanone (MEK)	27.94
2-Hexanone	31.41

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ISO 9001 QS Registered
ISO 17025-34-35-43 Accredited
Scopes: http://AbsoluteStandards.com

Run 12, "P82402 L070313 [2000µg/ml in M:W(9:1)]"

Run Length: 59.99 min, 35997 points at 10 points/second.
Created: Mon, Jul 8, 2013 at 10:21:12 PM.
Sampled: Sequence "070813-GC1", Method "GC1-M7".
Analyzed using default settings.

Comments
GC1-M7 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow = 150mL/min., Helium (carrier) = 10mL/min.,
Helium (make-up) = 40mL/min., Hydogen (make-up) = 100mL/min.,
Oven Profile: Temp. 1 = 35°C (Time 1 = 10 min.), Temp 2 = 200°C (Time 2 = 8.75 min.),
Rate = 4°C/min., Total run time = 60 min. Injector temp. = 200°C, PID Temp. = 200°C.
ELCD Signal = Edaq Channel 1 PID Signal = Edaq Channel 2
Standard injection = 0.5µL, Range=4 Purge Valve = 8 min



Name	PID RT (min.)
Acetone	8.69
2-Butanone	15.44
4-Methyl-2-pentanone	24.47
2-Hexanone	27.54

Absolute Standards, Inc.
800-368-1131
www.absolutestandards.com

Certified Reference Material CRM

ISO 9001:2008 Registered
ISO 17025-34 -43 Accredited
Scopes: http://AbsoluteStandards.com

Run 10, "P82402 L070313 [2000µg/mL in M:W(9:1)]"

Run Length: 59.99 min, 35995 points at 10 points/second.
Created: Mon, Jul 8, 2013 at 6:38:54 PM.
Sampled: Sequence "070813-GC5M1", Method "GC5-M1".
Analyzed using default settings.

Comments
GC5-M1 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow=290mL/min., Helium (carrier)=10mL/min.,
Helium(make-up)=10mL/min., Hydogen(make-up)=40mL/min., Air(make-up)=230mL/min.
Oven Profile: Temp. 1=35°C (Time 1=10 min.), Temp 2=200°C (Time 2=8.75 min.),
Rate = 4°C/min., Total run time=60 min. Injector temp.=200°C, FID Temp.=200°C.
FID Signal = Edaq Channel 1
Standard injection = 0.5µL, Range=3



Name	FID RT (min.)
Acetone	8.48
2-Butanone	15.29
4-Methyl-2-pentanone	24.95
2-Hexanone	28.30

125 Market Street
New Haven, CT 06513
USA

AccuStandard® Inc.

Tel (203)786-5290
Fax (203)786-5287
www.AccuStandard.com

Rec'd 9/05/12

CERTIFICATE OF ANALYSIS

Catalog No: APP-9-035-20X
Description: Carbon disulfide
Lot: 212071234
Solvent: Methanol

Date Certified: Jul 18, 2012
Expiration: Jul 18, 2022
Sample Size: 1 mL
Storage Condition: Refrig (0-5° C)
Hazards: HIGHLY FLAMMABLE

[X] Included on ISO/IEC 17025 Scope of Accreditation
[X] Included on ISO Guide 34 Scope of Accreditation

Component	CAS #	Purity % MFG	Prepared Concentration[1] (µg/mL)	Certified Analyte Concentration[2] (µg/mL)
Carbon disulfide	75-15-0	99.9	2005	2003

AccuStandard follows U.S. conventions in reporting numerical values on both certificates and labels:

A comma (,) is used to separate units of one-thousand or greater.
A period (.) is used as a decimal place marker.

1. All weights are traceable through NIST, Test No.822-275872-11
2. Certified Analyte Concentration = Purity x Prepared Concentration. The Uncertainty associated with the gravimetric values reported on this certificate is ±0.24%. The CRM Uncertainty calculated for this product is ±5%. These values are the expanded uncertainty and represent an estimated standard deviation equal to the positive square root of the total variation of the uncertainty of components. A normal distribution is assumed and a coverage factor of K=2 is chosen using approximately a 95% confidence level.
3. A product with a suffix (-1A, -2B, etc. or -01, -02, etc.) on its lot# has had its expiration date extended and is identical to the same lot# without the suffix.

For use in routine laboratory analysis.

See reverse side for additional information
Refer to the MSDS for additional safety information

Certified by: R. Cooper
Russ Cooper, QC Manager

AccuStandard is accredited to ISO Guide 34, ISO/IEC 17025 and certified to ISO 9001

OR-ORG/INO-001
Rev. 7/11

Certificate of Analysis

Rec'd 1/11/13

ULTRA SCIENTIFIC
Analytical Solutions

VOC Gas Mixture

Product Number: DWM-544 **Page:** 1 of 1

Lot Number: CJ-4469 **Lot Issue Date:** 18-Dec-2012 **Expiration Date:** 31-Jan-2016

This certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
bromomethane	000074-83-9	RM00064	2001 ± 10 µg/mL
chloroethane	000075-00-3	RM00065	2006 ± 10 µg/mL
chloromethane	000074-87-3	RM05290	2009 ± 10 µg/mL
dichlorodifluoromethane	000075-71-8	RM05289	2008 ± 10 µg/mL
trichlorofluoromethane	000075-69-4	RM00017	2005 ± 10 µg/mL
vinyl chloride	000075-01-4	RM05288	2008 ± 10 µg/mL

Matrix: methanol (methyl alcohol)

Storage: Store at < 4° C

ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.

A2LA ACCREDITED
ISO 17025:2005 Accredited A2LA Cert. No. 0851-01

ISO 9001:2008 Registered TUV USA, Inc. Cert. No. 09-1009

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
wwww.ultrasci.com

William J. Leary
Quality Assurance Manager

Rec'd 6/13/13

Methanol
ULTRA RESI-ANALYZED
For Purge and Trap Analysis

J.T.Baker | Part of AVANTOR™ PERFORMANCE MATERIALS

Material No.: 9077-02
Batch No.: 0000038701
Manufactured Date: 2013/02/20
Retest Date: 2018/02/19

Certificate of Analysis

Test	Specification	Result
Assay (CH_3OH) (by GC, corrected for water)	>= 99.9 %	100.0
Residue after Evaporation	<= 1.0000 ppm	< 0.1000
Titrable Acid (µeq/g)	<= 0.3	<0.1
Titrable Base (µeq/g)	<= 0.1	<0.01
Water (by KF, coulometric)	<= 0.08 %	<0.01
Photolonization Detection (PID) Below CRQL	Passes Test	PT
Electroconductivity Detection (ELCD) Below CRQL	Passes Test	PT

For Laboratory, Research or Manufacturing Use
Performance Tested for Use in EPA Methods
500 Series for Drinking Water
600 Series for Wastewater
846 for Solid Waste

Country of Origin: US
Packaging Site: Phillipsburg Mfg Ctr & DC

ISO

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.573.2600
Avantor ™ Performance Materials Inc.
3477 Corporate Parkway, Suite #200, Center Valley, PA 18034. U.S.A. Phone: 610.573.2600 . Fax: 610.573.2610

X1034 ISTD - only

80 ml STM-520-1 lot CH-3660 2500 ug/ml (R4399) exp. 12/31/14
3920 ml MeOH (R4493)
4000 ml 50 ug/ml LS 2/21/14 Exp. 2/28/14

X1035 Surrogate - only

80 ml STM-5301 lot CJ-1880 2500 ug/ml (R4392) exp. 5/31/15
3920 ml MeOH (R4493)
4000 ml 50 ug/ml LS 2/21/14 Exp. 2/28/14

X1036 ICV MIX

8260 Mix (ABS)	MTBE (S)	RMA (ACC)	Ketones (ABS)	CS2 (ACC)	MEOH	Final Vol.
32001 Lot 100113 2000 ug/mL R4733 Exp.10/01/16	CRM48483 Lot LB94566 2000 ug/mL R4517 Exp. 07/2015	S-2932-R1 Lot 212081711 Varied Conc. R4413 Exp. 9/04/14	82402 Lot 070313 2000 ug/mL R4511 Exp. 07/03/16	APP-9-035-20X Lot 212071234 2.0 mg/mL R4438 Exp. 07/18/22	R4493	(uL)
25	25	250	50	25	1625	2000

Ketones/Acrylo/
m&p-Xylenes/DBCP = 50 ug/mL
OTHER = 25 ug/mL

Prep: 02/21/14

EXP: 02/28/14

LS 2/21/14

X1037 Gases - Ultra

100 ml DWM-544 lot CJ-4469 2000 ug/ml (R4367) exp. 1/31/14
3900 ml MeOH (R4493)
4000 ml 50 ug/ml LS 2/21/14 Exp. 2/28/14
3/07/14 (RMA)

Continued on Page

Read and Understood By

Signed	Date	Signed	Date
LS	2/21/14	TRC	2/27/14

Rec'd 1/28/14

TIFIED WEIGHT REPORT

		Lot #	Solvent(s):
Part Number:	32001	DG596Q1	Methanol
Lot Number:	100113		
Description:	EPA Method 502.2/524.2 Volatile Organic Compounds 54 components		
Expiration Date:	100116		
Recommended Storage:	Freezer (0 °C)		
Nominal Concentration (µg/mL):	2000		

Formulated By:	Vincent K. Criscio, Jr.	100113	DATE
Reviewed By:	Pedro L. Rentas	100113	DATE

5E-05 Balance Uncertainty

Volume(s) shown below were combined and diluted to: 200.0 0.015 Flask Uncertainty

Compound	Part Number	Lot Number	Dil. Factor	Initial Vol. (mL)	Uncertainty Pipette	Initial Conc.(ug/mL)	Final Conc.(ug/mL)	Expanded Uncertainty	CAS#	MSDS Information (Solvent Safety Info. On Attached pg.) OSHA PEL (TWA)	LD50
1. Bromodichloromethane	30171	122112	0.10	20.00	0.011	20016.1	2001.5	0.00408	00075-27-4	N/A	orl-rat 916mg/kg
2. Dibromochloromethane	30171	122112	0.10	20.00	0.011	20011.7	2001.1	0.00408	00124-48-1	N/A	orl-rat 848mg/kg
3. cis-1,2-Dichloroethene	30171	122112	0.10	20.00	0.011	20018.4	2001.7	0.00404	00156-59-2	N/A	N/A
4. trans-1,2-Dichloroethene	30171	122112	0.10	20.00	0.011	20001.9	2000.1	0.00408	00156-60-5	N/A	orl-rat 1235mg/kg
5. Methylene chloride	30171	122112	0.10	20.00	0.011	20004.5	2000.4	0.00401	00075-09-2	500 ppm	orl-rat 2136mg/kg
6. 1,1-Dichloroethene	32251	100113	0.10	20.00	0.011	20009.8	2000.9	0.00404	00075-35-4	1 ppm (4mg/m3/8H)	orl-rat 200mg/kg
7. Bromochloromethane	30160	052313	0.10	20.00	0.011	20007.3	2000.6	0.00404	00074-97-5	200 ppm (1050mg/m3/8H)	orl-rat 5000mg/kg
8. Bromoform	30160	052313	0.10	20.00	0.011	20005.9	2000.5	0.00404	00075-25-2	0.5 ppm (5mg/m3) (skin)	orl-rat 1147mg/kg
9. Carbon tetrachloride	30160	052313	0.10	20.00	0.011	20005.6	2000.5	0.00402	00056-23-5	2 ppm (12.6mg/m3/8H)	orl-rat 2350mg/kg
10. Chloroform	30160	052313	0.10	20.00	0.011	20007.3	2000.6	0.00400	00067-66-3	50 ppm (240mg/m3) (CL)	orl-rat 908mg/kg
11. Dibromomethane	30160	052313	0.10	20.00	0.011	20006.1	2000.5	0.00404	00074-95-3	N/A	orl-rat 108mg/kg
12. 1,1-Dichloroethane	30160	052313	0.10	20.00	0.011	20007.3	2000.6	0.00404	00075-34-3	100 ppm	orl-rat 725mg/kg
13. 2,2-Dichloropropane	30160	052313	0.10	20.00	0.011	20005.9	2000.5	0.00408	00594-20-7	N/A	N/A
14. Tetrachloroethene	30160	052313	0.10	20.00	0.011	20008.0	2000.7	0.00404	00127-18-4	5 ppm (170mg/m3/8H)(fina	orl-rat 2629mg/kg
15. 1,1,1-Trichloroethane	30160	052313	0.10	20.00	0.011	20007.3	2000.6	0.00404	00071-55-6	350 ppm (1900mg/m3/8H)	orl-rat 10300mg/kg
16. 1,2-Dibromo-3-chloropropane	30161	072312	0.10	20.00	0.011	20020.9	2002.0	0.00404	00096-12-8	0.001 ppm	orl-rat 170mg/kg
17. 1,2-Dibromoethane	30161	072312	0.10	20.00	0.011	20023.4	2002.2	0.00404	00106-93-4	20 ppm (8H)	orl-rat 108mg/kg
18. 1,2-Dichloroethane	30161	072312	0.10	20.00	0.011	20004.5	2000.4	0.00401	00107-06-2	50 ppm (8H)	orl-rat 670mg/kg
19. 1,2-Dichloropropane	30161	072312	0.10	20.00	0.011	20022.9	2002.2	0.00404	00078-87-5	75 ppm (350mg/m3/8H)	orl-rat 1947mg/kg
1,3-Dichloropropane	30161	072312	0.10	20.00	0.011	20007.0	2000.6	0.00404	00142-28-9	N/A	unr-mus 3600mg/kg
,1-Dichloropropene	30161	072312	0.10	20.00	0.011	20022.3	2002.1	0.00404	00563-58-6	N/A	N/A
22. cis-1,3-Dichloropropene	30161	072312	0.10	20.00	0.011	20014.4	2001.3	0.00404	10061-01-5	N/A	N/A
23. trans-1,3-Dichloropropene	30161	072312	0.10	20.00	0.011	20012.2	2001.1	0.00404	10061-02-6	N/A	N/A
24. Hexachloro-1,3-butadiene	30161	072312	0.10	20.00	0.011	20007.3	2000.6	0.00408	00087-68-3	0.02 ppm (0.24mg/m3/8H)	orl-rat 90mg/kg
25. 1,1,1,2-Tetrachloroethane	30161	072312	0.10	20.00	0.011	20007.5	2000.7	0.00404	00630-20-6	N/A	orl-rat 670mg/kg
26. 1,1,2,2-Tetrachloroethane	30161	072312	0.10	20.00	0.011	20008.0	2000.7	0.00404	00079-34-5	5 ppm (35mg/m3/9H)(skin)	orl-rat 800mg/kg
27. 1,1,2-Trichloroethane	30161	072312	0.10	20.00	0.011	20008.2	2000.7	0.00408	00079-00-5	10 ppm (45mg/m3/8H)(skin	orl-rat 836mg/kg
28. Trichloroethene	30161	072312	0.10	20.00	0.011	20012.5	2001.2	0.00402	00079-01-6	50 ppm (270mg/m3/8H)	orl-mus 2402mg/kg
29. 1,2,3-Trichloropropane	30161	072312	0.10	20.00	0.011	20009.5	2000.9	0.00404	00096-18-4	10 ppm (60mg/m3/8H)	orl-rat 320mg/kg
30. Benzene	30162	071613	0.10	20.00	0.011	20006.4	2000.6	0.00401	00071-43-2	1 ppm	orl-rat 4894mg/kg
31. Bromobenzene	30162	071613	0.10	20.00	0.011	20008.2	2000.7	0.00404	00108-86-1	N/A	orl-rat 2699mg/kg
32. n-Butyl benzene	30162	071613	0.10	20.00	0.011	20014.4	2001.3	0.00404	00104-51-8	N/A	N/A
33. Ethyl benzene	30162	071613	0.10	20.00	0.011	20010.0	2000.9	0.00401	00100-41-4	100 ppm (435mg/m3/8H)	orl-rat >2000mg/kg
34. p-Isopropyl toluene	30162	071613	0.10	20.00	0.011	20012.0	2001.1	0.00404	00099-87-6	N/A	orl-rat 4750mg/kg
35. Naphthalene	30162	071613	0.10	20.00	0.011	20018.7	2001.8	0.00404	00091-20-3	10 ppm (50mg/m3/8H)	orl-rat 490mg/kg
36. Styrene	30162	071613	0.10	20.00	0.011	20019.9	2001.9	0.00404	00100-42-5	100 ppm	orl-rat 5000mg/kg
37. Toluene	30162	071613	0.10	20.00	0.011	20011.6	2001.1	0.00404	00108-88-3	200 ppm	orl-rat 5000mg/kg
38. 1,2,3-Trichlorobenzene	30162	071613	0.10	20.00	0.011	20020.0	2001.9	0.00404	00087-61-6	N/A	ipr-mus 1390mg/kg
39. 1,2,4-Trichlorobenzene	30162	071613	0.10	20.00	0.011	20013.4	2001.2	0.00404	00120-82-1	5 ppm (CL) (40mg/m3)	orl-rat 756mg/kg
40. 1,2,4-Trimethylbenzene	30162	071613	0.10	20.00	0.011	20012.6	2001.2	0.00408	00095-63-6	N/A	orl-rat 5g/kg
41. 1,3,5-Trimethylbenzene	30162	071613	0.10	20.00	0.011	20007.4	2000.7	0.00408	00108-67-8	N/A	N/A
42. m-Xylene	30162	071613	0.10	20.00	0.011	20010.8	2001.0	0.00404	00108-38-3	100 ppm (435mg/m3/8H)	orl-rat 5g/kg
43. tert-Butyl benzene	30163	071913	0.10	20.00	0.011	20010.9	2001.0	0.00404	00098-06-6	N/A	N/A
44. sec-Butyl benzene	30163	071913	0.10	20.00	0.011	20014.4	2001.3	0.00404	00135-98-8	N/A	orl-rat 2240mg/kg
45. Chlorobenzene	30163	071913	0.10	20.00	0.011	20011.0	2001.0	0.00404	00108-90-7	75 ppm (350mg/m3/8H)	orl-rat 2290mg/kg
46. 2-Chlorotoluene	30163	071913	0.10	20.00	0.011	20014.5	2001.4	0.00404	00095-49-8	50 ppm (250mg/m3/8H)	unr-rat 5700mg/kg
47. 4-Chlorotoluene	30163	071913	0.10	20.00	0.011	20009.4	2000.8	0.00408	00106-43-4	N/A	orl-rat 2100mg/kg
48. 1,2-Dichlorobenzene	30163	071913	0.10	20.00	0.011	20006.1	2000.5	0.00404	00095-50-1	50 ppm (300mg/m3) (CL)	orl-rat 500mg/kg
49. 1,3-Dichlorobenzene	30163	071913	0.10	20.00	0.011	20006.6	2000.6	0.00408	00541-73-1	N/A	ipr-mus 1062mg/kg
50. 1,4-Dichlorobenzene	30163	071913	0.10	20.00	0.011	20017.2	2001.6	0.00404	00106-46-7	75 ppm (450mg/m3/8H)	orl-rat 500mg/kg
51. Isopropylbenzene	30163	071913	0.10	20.00	0.011	20011.4	2001.0	0.00408	00098-82-8	50 ppm (245mg/m3/8H)	orl-rat 1400mg/kg
52. n-Propylbenzene	30163	071913	0.10	20.00	0.011	20010.0	2000.9	0.00408	00103-65-1	N/A	orl-rat 6040mg/kg
53. o-Xylene	30163	071913	0.10	20.00	0.011	20005.6	2000.5	0.00404	00095-47-6	100 ppm (435mg/m3/8H)	ipr-mus 1364mg/kg
n-Xylene	30163	071913	0.10	20.00	0.011	20013.8	2001.3	0.00404	00106-42-3	100 ppm (435mg/m3/8H)	orl-rat 5g/kg

Run 10, "P32001 L100113 [2000µg/mL in MeOH]-PID"

Run Length: 59.99 min, 35996 points at 10 points/second.
Created: Tue, Oct 1, 2013 at 8:42:31 PM.
Sampled: Sequence "100113-GC1", Method "GC1-M7".
Analyzed using Method "GC1-M7".

Comments
GC1-M7 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow = 150mL/min., Helium (carrier) = 10mL/min.,
Helium (make-up) = 40mL/min., Hydogen (make-up) = 100mL/min.,
Oven Profile: Temp. 1 = 35°C (Time 1 = 10 min.), Temp 2 = 200°C (Time 2 = 8.75 min.),
Rate = 4°C/min., Total run time = 60 min. Injector temp. = 200°C, PID Temp. = 200°C.
ELCD Signal = Edaq Channel 1 PID Signal = Edaq Channel 2
Standard Injection = 0.5µL, Range=4 Purge Valve = 8 min



Analyte	PID RT (min.)	ELCD RT (min.)
1,1-Dichloroethene	8.90	8.92
Methylene chloride	NA	10.87
trans-1,2-Dichloroethene	11.91	11.93
1,1-Dichloroethane	NA	13.68
2,2-Dichloropropane	NA	15.76
cis-1,2-Dichloroethene	15.98	16.00
Chloroform	NA	16.58
Bromochloromethane	NA	17.20
1,1,1-Trichloroethane	NA	17.95
1,1-Dichloropropene	18.50	18.52
Carbon tetrachloride	NA	18.82
1,2-Dichloroethane	NA	19.59
Benzene	19.50	NA
Trichloroethene	21.65	21.68
1,2-Dichloropropane	NA	22.37
Bromodichloromethane	NA	23.22
Dibromomethane	NA	23.41
cis-1,3-Dichloropropene	25.14	25.17
Toluene	26.16	NA
trans-1,3-Dichloropropene	26.95	26.97
1,1,2-Trichloroethane	NA	27.48
1,3-Dichloropropane	NA	28.39
Tetrachloroethene	28.36	28.39
Dibromochloromethane	NA	29.24
1,2-Dibromoethane	NA	29.92
Chlorobenzene	31.34	31.35
1,1,1,2-Tetrachloroethane	NA	31.54
Ethylbenzene	31.52	NA
m-Xylene	31.79	NA
p-Xylene	31.79	NA
o-Xylene	33.32	NA
Styrene	33.50	NA
Isopropylbenzene	34.61	NA
Bromoform	NA	34.73
1,1,2,2-Tetrachloroethane	NA	35.37
1,2,3-Trichloropropane	NA	35.89
n-Propylbenzene	36.05	NA
Bromobenzene	36.19	36.21
1,3,5-Trimethylbenzene	36.64	NA
2-Chlorotoluene	36.64	36.69
4-Chlorotoluene	36.85	36.87
tert-Butylbenzene	37.88	NA
1,2,4-Trimethylbenzene	38.02	NA
sec-Butylbenzene	38.64	NA
p-Isopropyltoluene	39.17	NA
1,3-Dichlorobenzene	39.48	39.50
1,4-Dichlorobenzene	39.88	39.91
n-Butylbenzene	40.65	NA
1,2-Dichlorobenzene	41.21	41.23
1,2-Dibromo-3-chloropropane	NA	44.18
1,2,4-Trichlorobenzene	47.30	47.32
Hexachlorobutadiene	47.80	47.82
Naphthalene	48.27	NA
1,2,3-Trichlorobenzene	49.17	49.19

Run 10, "P32001 L100113 [2000µg/mL in MeOH]-ELCD"

Run Length: 59.99 min, 35996 points at 10 points/second.
Created: Tue, Oct 1, 2013 at 8:42:31 PM.
Sampled: Sequence "100113-GC1", Method "GC1-M7".
Analyzed using Method "GC1-M7".

Comments
GC1-M7 Analysis by Candice Warren
Column ID SPB-Vocol 105 meter X 0.53mm X 3.0µm film thickness
Flow rates: Total flow = 150mL/min., Helium (carrier) = 10mL/min.,
Helium (make-up) = 40mL/min., Hydogen (make-up) = 100mL/min.,
Oven Profile: Temp. 1 = 35°C (Time 1 = 10 min.), Temp 2 = 200°C (Time 2 = 8.75 min.),
Rate = 4°C/min., Total run time = 60 min. Injector temp. = 200°C, PID Temp. = 200°C.
ELCD Signal = Edaq Channel 1 PID Signal = Edaq Channel 2
Standard Injection = 0.5µL, Range=4 Purge Valve = 8 min



X1038 IC MIX

8260 Mix (U)	MTBE (U)	RMA (CS)	Ketones (RES)	CS2 (ABS)		MEOH	Final Vol.
DWM-589N Lot CK-2561 2000 ug/mL R4502 Exp.09/30/16	STS-440-1 Lot CG-4444 2000 ug/mL R4230 Exp. 01/31/15	SP-1022-345DSZ Lot 2246900 Varied Conc. R4649 Exp. 12/31/14	30006 Lot A090981 5000 ug/mL R4531 Exp. 12/2015	70060 Lot 121813 1000 ug/mL R4661 Exp. 12/18/16		R4493	(uL)
200	200	400	160	400		2640	4000

Ketones = 200 ug/mL
Acrylo/m&p-Xylenes = 200ug/mL
All other = 100ug/ml

Prep: 02/24/14
Exp: 03/~~07~~/14 03 US 2/26/14

US 2/24/14

X1039 Gases - Abs

100mL 30058 lot 010213 2000 ug/mL (R4445) exp. 01/02/16
3900 mL MeOH (R4479)
4000mL 50 ug/mL US 2/24/14 Exp. 3/~~07~~ 03/14
US 2/24/14

X1040 Anti-Foam Agent (Purge Water)

300 mg 31822 lot A096340 Neat (R4535) exp. 7/2018 Restock
DI Water
1500 mL US 2/25/14 Exp. 2/25/15

X1041 MIX for Archon - UM85

40uL 8260 MIX(U) DWM-589N-1 lot CK-2561 2000 ug/mL (R4502) exp. 9/30/16
16uL DBCP(U) EPA-1021 lot CH-1351 5000 ug/mL (R4376) exp. 5/31/15
40uL CS2(Acc) ARP-9-035-20X lot 2207 1234 2.0 mg/mL (R4438) exp. 7/18/22
80uL RMA(CS) SP-1022345DSZ lot 2246900 Varied Conc. (R4649) exp. 12/31/14
1824uL MeOH (R4493)
2000uL Acrylo/DBCP = 80 ug/mL
Other = 40 ug/mL US 2/25/14 Exp. 3/11/14

Continued on Page

Read and Understood By

US 2/25/14 JRC 2/27/14
Signed Date Signed Date

X1038 IC MIX

8260 Mix (U)	MTBE (U)	RMA (CS)	Ketones (RES)	CS2 (ABS)		MEOH	Final Vol.
DWM-589N Lot CK-2561 2000 ug/mL R4502 Exp.09/30/16	STS-440-1 Lot CG-4444 2000 ug/mL R4230 Exp. 01/31/15	SP-1022-345DSZ Lot 2246900 Varied Conc. R4649 Exp. 12/31/14	30006 Lot A090981 5000 ug/mL R4531 Exp. 12/2015	70060 Lot 121813 1000 ug/mL R4661 Exp. 12/18/16		R4493	(uL)
200	200	400	160	400		2640	4000

Ketones = 200 ug/mL
Acrylo/m&p-Xylenes = 200ug/mL
All other = 100ug/ml

Prep: 02/24/14
Exp: 03/~~07~~/14 03 US 2/26/14

US 2/24/14

X1039 Gases - Abs

100uL 30058 lot 010213 2000 ug/mL (R4445) exp. 01/08/16
3900 uL MeOH (R4479)
4000uL 50 ug/mL US 2/24/14 Exp. 3/~~07~~ 03/14 US 2/24/14

X1040 Anti-Foam Agent (Purge Water)

300 mg 31822 lot A0916340 Neat (R4535) exp. 7/2018 Restck
DI Water
1500 mL US 2/25/14 Exp. 2/25/15

X1041 MIX for Archon - UM85

40uL 8260 MIX(U) DWM-589N-1 lot CK-2561 2000 ug/mL (R4502) exp. 9/30/16
16uL DBCP(U) EPA-1021 lot CH-1351 5000 ug/mL (R4376) exp. 5/31/15
40uL CS2(Acc) APP-9-035-20X lot 2120 71234 2.0 mg/mL (R4438) exp. 7/18/22
80uL RMA(CS) SP-1022345DSZ lot 2246900 Varied Conc. (R4649) exp. 12/31/14
1824uL MeOH (R4493)
2000uL Acrylo/DBCP = 80 ug/mL
Other = 40 ug/mL US 2/25/14 Exp. 3/11/14

Continued on Page

Read and Understood By

US 2/25/14 TRC 2/27/14
Signed Date Signed Date

X1047 RMA Extracs - Accu Std

250 uL S-2932-R1 lot 212081711 Varied Conc. (R4413) exp. 4/30/15
750 uL MeOH (R4493)
1000 uL 50 ug/mL LS 2/28/14 Exp. 3/14/14

X1048 Ketones / CS_2 - Ultra

250 uL CMS-1707 lot CK-1031 Varied Conc (R44461) exp. 4/30/15
750 uL MeOH (R4493)
1000 uL Acetone / MEK = 250 ug/mL
MIBK / MNBK = 100 ug/mL
CS_2 = 50 ug/mL LS 2/28/14 Exp. 3/14/14

X1049 ISTD - mix

80 uL STM-520-1 lot CH-3660 2500 ug/mL (R4399) exp. 12/31/14
3920 uL MeOH (R4493)
4000 uL 50 ug/mL LS 3/03/14 Exp. 3/10/14

X1050 Surrogate - mix

80 uL STM-530-1 lot CJ-1280 2500 ug/mL (R4392) exp. 5/31/15
3920 uL MeOH (R4493)
4000 uL 50 ug/mL LS 3/03/14 Exp. 3/10/14

X1051 IEV Mix

8260 Mix (ABS)	MTBE (S)	RMA (ACC)	Ketones (ABS)	CS2 (ACC)	MEOH	Final Vol. (uL)
32001 Lot 100113 2000 ug/mL R4733 Exp.10/01/16	CRM48483 Lot LB94566 2000 ug/mL R4517 Exp. 07/2015	S-2932-R1 Lot 212081711 Varied Conc. R4413 Exp. 9/04/14	82402 Lot 070313 2000 ug/mL R4511 Exp. 07/03/16	APP-9-035-20X Lot 212071234 2.0 mg/mL R4438 Exp. 07/18/22	R4493	
25	25	250	50	25	1625	2000

Ketones/Acrylo/
m&p-Xylenes/DBCP = 50 ug/mL
OTHER = 25 ug/mL

Prep: 03/03/14
EXP: 03/10/14

LS 3/03/14

Continued on Page

Read and Understood By

LS 3/03/14 Signed Date

[illegible] 3/3/14 Signed Date

Analytical Run Number 1

Calibration Information

Tune Data (If applicable)

Initial Calibration (If different than Initial Calibration in Run 1)
- Calibration Factor Summary
- Chromatogram
- Quant Report (If applicable)

Continuing Calibration
- Calibration Response Check
- Chromatogram
- Quant Report (If applicable)

Instrument Logs/Analysis Notebook Pages

Blank / Field Sample / QC Data

Sample Results
Chromatogram
Quant Report (If applicable)
Confirmation Analysis Quant Report / Chromatogram (If applicable)
Tentatively Identified Compounds (If required)

Rev. 1.1

Calibration Information

Tune Data (If applicable)

Initial Calibration (If different than Initial Calibration in Run 1)
- Calibration Factor Summary
- Chromatogram
- Quant Report (If applicable)

Continuing Calibration
- Calibration Response Check
- Chromatogram
- Quant Report (If applicable)

Rev. 1.1

5A

VOLATILE ORGANIC GC/MS TUNING AND MASS CALIBRATION - BROMOFLUOROBENZENE (BFB)

Lab Name: ARDL, Inc. Contract: ---

Lab Code: --- Case No.: 006611 SAS No.: --- SDG No.: ---

Lab File ID: A0212408 BFB Injection Date: 02/12/14

Instrument ID: AG1 BFB Injection Time: 14:43

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE	
50	8.0 - 40.0% of mass 95	17.1	
75	30.0 - 66.0% of mass 95	46.9	
95	Base Peak, 100% relative abundance	100.	
96	5.0 - 9.0% of mass 95	7.4	
173	Less than 2.0% of mass 174	0.0	(1)
174	Greater than 50.0% of mass 95	73.7	
175	4.0 - 9.0% of mass 174	8.0	(1)
176	Greater than 93.0%, but less that 101.0% of mass 174	100.8	(1)
177	5.0 - 9.0% of mass 176	6.1	(2)

(1)-Value is % mass 174 2-Value is % mass 176

THIS TUNE APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
1	BLK STD	000 STD	A0212411	02/12/14	16:06
2	VOA STD	12.5 STD	A0212412	02/12/14	16:36
3	VOA STD	025 STD	A0212413	02/12/14	17:04
4	VOA STD	050 STD	A0212414	02/12/14	17:33
5	VOA STD	075 STD	A0212415	02/12/14	18:03
6	VOA STD	100 STD	A0212416	02/12/14	18:32
7	VOA STD	125 STD	A0212417	02/12/14	19:02
8	VOA STD	150 STD	A0212418	02/12/14	19:32
9	VOA STD	200 STD	A0212419	02/12/14	20:03
10					
11					
12					
13					
14					
15					
16					
17					
18					
19					
20					
21					
22					

Data Path : C:\MSDCHEM\1\DATA\021214\
Data File : A0212408.D
Acq On : 12 Feb 2014 2:43 pm
Operator :
Sample : 50 NG BFB
Misc : 8260
ALS Vial : 1 Sample Multiplier: 1

Integration File: rteint.p

Method : C:\MSDCHEM\1\METHODS\BFBSOIL.M
Title :
Last Update :





AutoFind: Scans 499, 500, 501; Background Corrected with Scan 490

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	17.1	11491	PASS
75	95	30	66	46.9	31506	PASS
95	95	100	100	100.0	67170	PASS
96	95	5	9	7.4	4947	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	73.7	49517	PASS
175	174	4	9	8.0	3969	PASS
176	174	93	101	100.8	49920	PASS
177	176	5	9	6.1	3069	PASS

Average of 4.683 to 4.689 min.: A0212408.D
50 NG BFB
Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
36.00	186	51.05	3475	70.00	222	93.00	2860
36.20	188	56.00	376	73.05	2609	94.00	6957
37.10	2609	56.20	351	74.05	9265	95.00	67170
38.05	2082	57.05	1345	75.05	31506	96.05	4947
39.05	990	60.05	672	76.10	2689	140.95	731
44.00	396	61.05	2763	76.90	190	142.80	254
45.00	481	62.10	2839	78.95	978	143.00	220
47.10	562	63.05	2070	80.95	1238	173.95	49517
48.20	173	64.00	167	87.00	2406	174.95	3969
49.00	2452	68.00	5774	87.95	2558	176.00	49920
50.10	11491	69.05	5444	92.00	1935	176.95	3069

Response Factor Report AG1

Method Path : C:\MSDCHEM\1\METHODS\
Method File : SOIL_44HI.M
Title : 8260 VOLATILES
Last Update : Thu Mar 20 08:47:48 2014
Response Via : Initial Calibration

Calibration Files
12.5=A0212412.D 25 =A0212413.D 50 =A0212414.D
75 =A0212415.D 100 =A0212416.D 125 =A0212417.D

		Compound	12.5	25	50	75	100	125	Avg	%RSD
1)	I	Chlorobenzene-d5	----------ISTD----------							
2)	S	Toluene-d8	1.834	1.970	2.327	2.462	2.350	2.487	2.266	10.36
3)	MCT	Toluene	2.260	3.369	3.117	3.069	2.947	3.329	3.029	11.43
4)	CT	Ethylbenzene	3.093	4.137	3.732	3.610	3.473	3.945	3.654	8.75
5)	T	m & p-Xylene	2.375	3.436	3.028	2.949	2.832	3.202	2.960	10.45
6)	T	Styrene	1.482	2.472	2.280	2.236	2.201	2.469	2.217	14.32
7)	T	o-Xylene	0.875	1.378	1.288	1.276	1.262	1.446	1.277	13.80
8)	I	1,4-Dichlorobenzene-d	----------ISTD----------							
9)	S	4-Bromofluorobenz	0.832	0.829	0.948	0.993	0.977	1.024	0.951	8.29
10)	T	1,2,4-Trimethylbe	2.402	3.578	3.438	3.346	3.340	3.719	3.351	12.10

(#) = Out of Range ### Number of calibration levels exceeded format ###

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\021214\
Data File : A0212411.D
Acq On : 12 Feb 2014 4:06 pm
Operator :
Sample : VSTD000
Misc : 8260
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Mar 20 08:23:36 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 08:23:31 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Chlorobenzene-d5	8.77	82	299125	125.00 ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	274646	125.00 ug/Kg	0.21

System Monitoring Compounds

2) Toluene-d8	0.00	98	0	0.00 ug/Kg	
Spiked Amount 100.000			Recovery	=	0.00%
9) 4-Bromofluorobenzene	0.00	95	0	0.00 ug/Kg	
Spiked Amount 100.000			Recovery	=	0.00%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\021214\
Data File : A0212411.D
Acq On : 12 Feb 2014 4:06 pm
Operator :
Sample : VSTD000
Misc : 8260
ALS Vial : 3 Sample Multiplier: 1

Quant Time: Mar 20 08:23:36 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 08:23:31 2014
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\021214\
Data File : A0212412.D
Acq On : 12 Feb 2014 4:36 pm
Operator :
Sample : VSTD12.5
Misc : 8260
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Mar 20 08:29:38 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 08:29:00 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Chlorobenzene-d5	8.77	82	250789	125.00 ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	265115	125.00 ug/Kg	0.00
System Monitoring Compounds					
2) Toluene-d8	7.83	98	45998	10.17 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 10.17%	
9) 4-Bromofluorobenzene	9.48	95	22046	11.39 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 11.39%	
Target Compounds					Qvalue
3) Toluene	7.87	91	56679	9.61 ug/Kg	99
4) Ethylbenzene	8.78	91	77557	11.14 ug/Kg	94
5) m & p-Xylene	8.86	91	119133	21.09 ug/Kg	93
6) Styrene	9.15	104	37160	8.82 ug/Kg	95
7) o-Xylene	9.12	106	21948	8.29 ug/Kg	87
10) 1,2,4-Trimethylbenzene	9.88	105	63676	9.43 ug/Kg	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\021214\
Data File : A0212412.D
Acq On : 12 Feb 2014 4:36 pm
Operator :
Sample : VSTD12.5
Misc : 8260
ALS Vial : 4 Sample Multiplier: 1

Quant Time: Mar 20 08:29:38 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 08:29:00 2014
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\021214\
Data File : A0212413.D
Acq On : 12 Feb 2014 5:04 pm
Operator :
Sample : VSTD025
Misc : 8260
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Mar 20 08:30:25 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 08:29:00 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Chlorobenzene-d5	8.77	82	249286	125.00 ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	266176	125.00 ug/Kg	0.00
System Monitoring Compounds					
2) Toluene-d8	7.83	98	98242	21.85 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 21.85%	
9) 4-Bromofluorobenzene	9.48	95	44133	22.72 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 22.72%	
Target Compounds					Qvalue
3) Toluene	7.86	91	167991	28.66 ug/Kg	99
4) Ethylbenzene	8.78	91	206260	29.81 ug/Kg	98
5) m & p-Xylene	8.86	91	342665	61.02 ug/Kg	94
6) Styrene	9.15	104	123233	29.43 ug/Kg	94
7) o-Xylene	9.12	106	68699	26.10 ug/Kg	89
10) 1,2,4-Trimethylbenzene	9.88	105	190500	28.10 ug/Kg	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\021214\
Data File : A0212413.D
Acq On : 12 Feb 2014 5:04 pm
Operator :
Sample : VSTD025
Misc : 8260
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Mar 20 08:30:25 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 08:29:00 2014
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\021214\
Data File : A0212414.D
Acq On : 12 Feb 2014 5:33 pm
Operator :
Sample : VSTD050
Misc : 8260
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Mar 20 08:31:01 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 08:29:00 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Chlorobenzene-d5	8.77	82	282745	125.00 ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	284885	125.00 ug/Kg	0.00
System Monitoring Compounds					
2) Toluene-d8	7.83	98	263178	51.60 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 51.60%	
9) 4-Bromofluorobenzene	9.48	95	108054	51.97 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 51.97%	
Target Compounds					Qvalue
3) Toluene	7.86	91	352532	53.03 ug/Kg	100
4) Ethylbenzene	8.78	91	422118	53.79 ug/Kg	97
5) m & p-Xylene	8.86	91	685030	107.55 ug/Kg	93
6) Styrene	9.15	104	257893	54.30 ug/Kg	94
7) o-Xylene	9.11	106	145708	48.82 ug/Kg	89
10) 1,2,4-Trimethylbenzene	9.88	105	391726	53.98 ug/Kg	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\021214\
Data File : A0212414.D
Acq On : 12 Feb 2014 5:33 pm
Operator :
Sample : VSTD050
Misc : 8260
ALS Vial : 6 Sample Multiplier: 1

Quant Time: Mar 20 08:31:01 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 08:29:00 2014
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\021214\
Data File : A0212415.D
Acq On : 12 Feb 2014 6:03 pm
Operator :
Sample : VSTD075
Misc : 8260
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Mar 20 08:31:44 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 08:29:00 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Chlorobenzene-d5	8.76	82	257326	125.00 ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	257061	125.00 ug/Kg	0.00
System Monitoring Compounds					
2) Toluene-d8	7.83	98	380092	81.89 ug/Kg	0.00
Spiked Amount 100.000			Recovery =	81.89%	
9) 4-Bromofluorobenzene	9.48	95	153123	81.62 ug/Kg	0.00
Spiked Amount 100.000			Recovery =	81.62%	
Target Compounds					Qvalue
3) Toluene	7.86	91	473913	78.33 ug/Kg	100
4) Ethylbenzene	8.78	91	557331	78.04 ug/Kg	96
5) m & p-Xylene	8.86	91	910493	157.06 ug/Kg	93
6) Styrene	9.15	104	345231	79.87 ug/Kg	93
7) o-Xylene	9.11	106	196996	72.52 ug/Kg	84
10) 1,2,4-Trimethylbenzene	9.88	105	516084	78.81 ug/Kg	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\021214\
Data File : A0212415.D
Acq On : 12 Feb 2014 6:03 pm
Operator :
Sample : VSTD075
Misc : 8260
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Mar 20 08:31:44 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 08:29:00 2014
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\021214\
Data File : A0212416.D
Acq On : 12 Feb 2014 6:32 pm
Operator :
Sample : VSTD100
Misc : 8260
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Mar 20 08:32:11 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 08:29:00 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Chlorobenzene-d5	8.76	82	279384	125.00 ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	268227	125.00 ug/Kg	0.00
System Monitoring Compounds					
2) Toluene-d8	7.83	98	525140	104.21 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 104.21%	
9) 4-Bromofluorobenzene	9.47	95	209623	107.08 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 107.08%	
Target Compounds					Qvalue
3) Toluene	7.86	91	658595	100.26 ug/Kg	100
4) Ethylbenzene	8.78	91	776236	100.11 ug/Kg	97
5) m & p-Xylene	8.86	91	1266059	201.15 ug/Kg	93
6) Styrene	9.15	104	491861	104.80 ug/Kg	91
7) o-Xylene	9.11	106	282160	95.67 ug/Kg	84
10) 1,2,4-Trimethylbenzene	9.88	105	716744	104.90 ug/Kg	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\021214\
Data File : A0212416.D
Acq On : 12 Feb 2014 6:32 pm
Operator :
Sample : VSTD100
Misc : 8260
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Mar 20 08:32:11 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 08:29:00 2014
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\021214\
Data File : A0212417.D
Acq On : 12 Feb 2014 7:02 pm
Operator :
Sample : VSTD125
Misc : 8260
ALS Vial : 9 Sample Multiplier: 1

Quant Time: Mar 20 08:39:39 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 08:29:00 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Chlorobenzene-d5	8.76	82	274744	125.00 ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	270361	125.00 ug/Kg	0.00
System Monitoring Compounds					
2) Toluene-d8	7.83	98	683333	137.89 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 137.89%	
9) 4-Bromofluorobenzene	9.47	95	276947	140.35 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 140.35%	
Target Compounds					Qvalue
3) Toluene	7.86	91	914534	141.58 ug/Kg	100
4) Ethylbenzene	8.78	91	1083877	142.15 ug/Kg	97
5) m & p-Xylene	8.86	91	1759378	284.25 ug/Kg	93
6) Styrene	9.14	104	678335	146.98 ug/Kg	93
7) o-Xylene	9.11	106	397346	137.00 ug/Kg	85
10) 1,2,4-Trimethylbenzene	9.87	105	1005436	145.99 ug/Kg	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\021214\
Data File : A0212417.D
Acq On : 12 Feb 2014 7:02 pm
Operator :
Sample : VSTD125
Misc : 8260
ALS Vial : 9 Sample Multiplier: 1

Quant Time: Mar 20 08:39:39 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 08:29:00 2014
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\021214\
Data File : A0212418.D
Acq On : 12 Feb 2014 7:32 pm
Operator :
Sample : VSTD150
Misc : 8260
ALS Vial : 10 Sample Multiplier: 1

Quant Time: Mar 20 08:40:01 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 08:29:00 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Chlorobenzene-d5	8.76	82	287555	125.00	ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	272237	125.00	ug/Kg	0.00
System Monitoring Compounds						
2) Toluene-d8	7.83	98	802593	154.74	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	154.74%	
9) 4-Bromofluorobenzene	9.47	95	320452	161.28	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	161.28%	
Target Compounds						Qvalue
3) Toluene	7.86	91	1022611	151.25	ug/Kg	100
4) Ethylbenzene	8.78	91	1201084	150.50	ug/Kg	97
5) m & p-Xylene	8.86	91	1953492	301.56	ug/Kg	92
6) Styrene	9.14	104	760052	157.35	ug/Kg	91
7) o-Xylene	9.11	106	443350	146.05	ug/Kg	83
10) 1,2,4-Trimethylbenzene	9.87	105	1112770	160.46	ug/Kg	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\021214\
Data File : A0212418.D
Acq On : 12 Feb 2014 7:32 pm
Operator :
Sample : VSTD150
Misc : 8260
ALS Vial : 10 Sample Multiplier: 1

Quant Time: Mar 20 08:40:01 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 08:29:00 2014
Response via : Initial Calibration



Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\021214\
Data File : A0212419.D
Acq On : 12 Feb 2014 8:03 pm
Operator :
Sample : VSTD200
Misc : 8260
ALS Vial : 11 Sample Multiplier: 1

Quant Time: Mar 20 08:40:34 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 08:29:00 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Chlorobenzene-d5	8.76	82	277381	125.00 ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	270685	125.00 ug/Kg	0.00
System Monitoring Compounds					
2) Toluene-d8	7.83	98	1051307	210.12 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 210.12%	
9) 4-Bromofluorobenzene	9.47	95	444587	225.04 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 225.04%	
Target Compounds					Qvalue
3) Toluene	7.86	91	1409514	216.13 ug/Kg	100
4) Ethylbenzene	8.77	91	1668416	216.73 ug/Kg	96
5) m & p-Xylene	8.86	91	2686892	429.98 ug/Kg	93
6) Styrene	9.14	104	1063677	228.28 ug/Kg	91
7) o-Xylene	9.11	106	624070	213.12 ug/Kg	85
10) 1,2,4-Trimethylbenzene	9.87	105	1550793	224.90 ug/Kg	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\021214\
Data File : A0212419.D
Acq On : 12 Feb 2014 8:03 pm
Operator :
Sample : VSTD200
Misc : 8260
ALS Vial : 11 Sample Multiplier: 1

Quant Time: Mar 20 08:40:34 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 08:29:00 2014
Response via : Initial Calibration



5A

VOLATILE ORGANIC GC/MS TUNING AND MASS CALIBRATION - BROMOFLUOROBENZENE (BFB)

Lab Name: ARDL, Inc. Contract: ---

Lab Code: --- Case No.: 006611 SAS No.: --- SDG No.: ---

Lab File ID: A0213402 BFB Injection Date: 02/13/14

Instrument ID: AG1 BFB Injection Time: 10:53

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE	
50	8.0 - 40.0% of mass 95	17.2	
75	30.0 - 66.0% of mass 95	48.9	
95	Base Peak, 100% relative abundance	100.	
96	5.0 - 9.0% of mass 95	6.7	
173	Less than 2.0% of mass 174	0.0	(1)
174	Greater than 50.0% of mass 95	73.2	
175	4.0 - 9.0% of mass 174	7.1	(1)
176	Greater than 93.0%, but less that 101.0% of mass 174	93.7	(1)
177	5.0 - 9.0% of mass 176	6.3	(2)

(1)-Value is % mass 174 2-Value is % mass 176

THIS TUNE APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
1	ICV	100 STD	A0213404	02/13/14	12:46
2					
3					
4					
5					
6					
7					
8					
9					
10					
11					
12					
13					
14					
15					
16					
17					
18					
19					
20					
21					
22					

BFB

4 us 02/26/14

Data Path : C:\MSDCHEM\1\DATA\021314\
Data File : A0213302.D
Acq On : 13 Feb 2014 10:53 am
Operator :
Sample : 50 NG BFB
Misc : 8260
ALS Vial : 1 Sample Multiplier: 1

Integration File: rteint.p

Method : C:\MSDCHEM\1\METHODS\BFB.M
Title :
Last Update :





AutoFind: Scans 499, 500, 501; Background Corrected with Scan 492

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	17.2	3951	PASS
75	95	30	66	48.9	11210	PASS
95	95	100	100	100.0	22906	PASS
96	95	5	9	6.7	1524	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	73.2	16756	PASS
175	174	4	9	7.1	1196	PASS
176	174	93	101	93.7	15700	PASS
177	176	5	9	6.3	995	PASS

cage of 4.683 to 4.690 min.: A0213302.D
NG BFB
dified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
37.05	914	73.00	896	95.00	22906		
38.10	804	74.05	3520	96.05	1524		
49.05	852	75.00	11210	173.95	16756		
50.05	3951	76.00	1102	174.95	1196		
51.05	1400	78.90	221	175.95	15700		
57.05	599	80.90	181	176.95	995		
61.05	1100	86.95	769				
62.05	961	88.00	518				
63.05	829	92.05	458				
68.00	2268	93.00	989				
69.00	2069	94.05	2562				

Continuing Calibration Report AG1

Method Path : C:\MSDCHEM\1\METHODS\
Method File : SOIL_44HI.M
Title : 8260 VOLATILES
Last Update : Thu Mar 20 09:21:05 2014
Response Via : Initial Calibration

CC Data File: A0213~~3~~4.D
4
3/20/14

Min. RRF : 0.000 Min. Rel. Area : 50%
Max. RRF Dev : 20% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%
1	I	Chlorobenzene-d5	1.000	1.000	0.0	97
2	S	Toluene-d8	2.266	2.428	-7.2	100
3	MCT	Toluene	3.029	3.142	-3.7	103
4	CT	Ethylbenzene	3.654	3.676	-0.6	102
5	T	m & p-Xylene	2.960	3.083	-4.2	105
6	T	Styrene	2.217	2.314	-4.3	102
7	T	o-Xylene	1.277	1.396	-9.3	107
8	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101
9	S	4-Bromofluorobenzene	0.951	1.002	-5.4	103
10	T	1,2,4-Trimethylbenzene	3.351	3.744	-11.7	113

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\021314\
Data File : A0213304.D
Acq On : 13 Feb 2014 12:46 pm
Operator :
Sample : VSTD100/ICV
Misc : 8260
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 20 09:20:58 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 09:11:30 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Chlorobenzene-d5	8.77	82	270393	125.00 ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	270377	125.00 ug/Kg	0.00
System Monitoring Compounds					
2) Toluene-d8	7.83	98	525248	107.18 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 107.18%	
9) 4-Bromofluorobenzene	9.48	95	216816	105.37 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 105.37%	
Target Compounds					Qvalue
3) Toluene	7.86	91	679686	103.74 ug/Kg	99
4) Ethylbenzene	8.78	91	795240	100.62 ug/Kg	97
5) m & p-Xylene	8.86	91	1333941	208.33 ug/Kg	94
6) Styrene	9.15	104	500492	104.35 ug/Kg	91
7) o-Xylene	9.12	106	301959	109.30 ug/Kg	85
10) 1,2,4-Trimethylbenzene	9.88	105	809782	111.71 ug/Kg	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\021314\
Data File : A0213304.D
Acq On : 13 Feb 2014 12:46 pm
Operator :
Sample : VSTD100/ICV
Misc : 8260
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 20 09:20:58 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 09:11:30 2014
Response via : Initial Calibration



SOIL_44HI.M Thu Mar 20 09:22:05 2014 AG1 Page: 2

5A

VOLATILE ORGANIC GC/MS TUNING AND MASS CALIBRATION - BROMOFLUOROBENZENE (BFB)

Lab Name: ARDL, Inc. Contract: ---

Lab Code: --- Case No.: 006613 SAS No.: --- SDG No.: ---

Lab File ID: A0227403 BFB Injection Date: 02/27/14

Instrument ID: AG1 BFB Injection Time: 09:10

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE	
50	8.0 - 40.0% of mass 95	18.8	
75	30.0 - 66.0% of mass 95	47.2	
95	Base Peak, 100% relative abundance	100.	
96	5.0 - 9.0% of mass 95	7.2	
173	Less than 2.0% of mass 174	0.0	(1)
174	Greater than 50.0% of mass 95	77.6	
175	4.0 - 9.0% of mass 174	8.1	(1)
176	Greater than 93.0%, but less that 101.0% of mass 174	98.4	(1)
177	5.0 - 9.0% of mass 176	7.3	(2)

(1)-Value is % mass 174 2-Value is % mass 176

THIS TUNE APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
1	VOA STD	100 STD	A0227404	02/27/14	09:42
2	SPBK 02/27/14	006613-04K1	A0227405	02/27/14	10:17
3	BLANK 02/27/14	006613-04B1	A0227406	02/27/14	10:53
4	MAK106L	006613-04	A0227413	02/27/14	14:29
5	MAK108,109,110,113L	006613-05	A0227414	02/27/14	14:57
6	MAK111L	006613-06	A0227415	02/27/14	15:25
7					
8					
9					
10					
11					
12					
13					
14					
15					
16					
17					
18					
19					
20					
21					
22					

Data Path : C:\MSDCHEM\1\DATA\022714\
Data File : A0227403.D
Acq On : 27 Feb 2014 9:10 am
Operator :
Sample : 50 NG BFB
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1

Integration File: rteint.p

Method : C:\MSDCHEM\1\METHODS\BFBSOIL.M
Title :
Last Update :





AutoFind: Scans 500, 501, 502; Background Corrected with Scan 492

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	18.8	5630	PASS
75	95	30	66	47.2	14126	PASS
95	95	100	100	100.0	29901	PASS
96	95	5	9	7.2	2155	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	77.6	23208	PASS
175	174	4	9	8.1	1877	PASS
176	174	93	101	98.4	22832	PASS
177	176	5	9	7.3	1664	PASS

Average of 4.686 to 4.693 min.: A0227403.D
50 NG BFB
Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
37.05	1063	68.05	2960	91.95	850		
38.05	1188	69.05	3078	92.95	1375		
39.05	428	73.00	1297	94.05	3589		
40.00	72	74.05	4851	95.05	29901		
49.05	1188	75.10	14126	96.05	2155		
50.05	5630	76.05	1158	143.00	170		
51.05	1798	78.95	686	174.00	23208		
57.05	838	80.95	803	174.95	1877		
61.05	1071	86.95	1137	176.00	22832		
62.05	1389	87.85	457	176.95	1664		
63.00	1012	88.10	348	207.10	196		

Continuing Calibration Report AG1

Method Path : C:\MSDCHEM\1\METHODS\
Method File : SOIL_44HI.M
Title : 8260 VOLATILES
Last Update : Thu Mar 20 10:27:26 2014
Response Via : Initial Calibration

CC Data File: A0227404.D

Min. RRF : 0.000 Min. Rel. Area : 50%
Max. RRF Dev : 20% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%
1	I	Chlorobenzene-d5	1.000	1.000	0.0	82
2	S	Toluene-d8	2.266	2.203	2.8	77
3	MCT	Toluene	3.029	3.042	-0.4	85
4	CT	Ethylbenzene	3.654	3.789	-3.7	89
5	T	m & p-Xylene	2.960	3.158	-6.7	91
6	T	Styrene	2.217	2.356	-6.3	88
7	T	o-Xylene	1.277	1.322	-3.5	86
8	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87
9	S	4-Bromofluorobenzene	0.951	0.918	3.5	82
10	T	1,2,4-Trimethylbenzene	3.351	3.552	-6.0	93

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022714\
Data File : A0227404.D
Acq On : 27 Feb 2014 9:42 am
Operator :
Sample : VSTD100
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 20 10:27:22 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:24:39 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Chlorobenzene-d5	8.77	82	228944	125.00	ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	233887	125.00	ug/Kg	0.00
System Monitoring Compounds						
2) Toluene-d8	7.83	98	403467	97.23	ug/Kg	0.00
Spiked Amount 100.000			Recovery = 97.23%			
9) 4-Bromofluorobenzene	9.48	95	171850	96.55	ug/Kg	0.00
Spiked Amount 100.000			Recovery = 96.55%			
Target Compounds						Qvalue
3) Toluene	7.86	91	557172	100.44	ug/Kg	99
4) Ethylbenzene	8.78	91	693962	103.70	ug/Kg	96
5) m & p-Xylene	8.86	91	1156851	213.38	ug/Kg	92
6) Styrene	9.15	104	431527	106.26	ug/Kg	88
7) o-Xylene	9.12	106	242159	103.53	ug/Kg	80
10) 1,2,4-Trimethylbenzene	9.88	105	664593	105.99	ug/Kg	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022714\
Data File : A0227404.D
Acq On : 27 Feb 2014 9:42 am
Operator :
Sample : VSTD100
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 20 10:27:22 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:24:39 2014
Response via : Initial Calibration



Instrument Logs/
Analysis Notebook Pages

Rev. 1.1

54
PROJECT 2/12/14 AGI SOIL

Notebook No. 3317
Continued from Page

Run Method = SOIL_24 M
Tune = BFBSOIL.U EM = 1887
Tekmar = NB 3167 pg. 34
Quant Method = SOIL_426

8	1	A0212408.d	0.	50 NG BFB (2ul - X1007) Pass	8260	
9	1	A0212409.d	0.	VSTD000	8260	
10	2	A0212410.d	0.	VSTD000	8260	Conc for SOIL_44HF
11	3	A0212411.d	0.	VSTD000 - Pass.	8260	
12	4	A0212412.d	0.	VSTD005	8260	- VSTD12.5
13	5	A0212413.d	0.	VSTD010	8260	- VSTD025
14	6	A0212414.d	0.	VSTD020	8260	- VSTD050
15	7	A0212415.d	0.	VSTD030	8260	- VSTD075
16	8	A0212416.d	0.	VSTD040 SOIL_43	8260	- VSTD100
17	9	A0212417.d	0.	VSTD050	8260	- VSTD125
18	10	A0212418.d	0.	VSTD060	8260	- VSTD150
19	11	A0212419.d	0.	VSTD080	8260	- VSTD200
20	12	A0212420.d	0.	FLUSH	8260	
21	13	A0212421.d	0.	VSTD005 Not Needed	8260	
22	14	A0212422.d	0.	VSTD030 ⊥	8260	

US 2/26/14

	ISTD X1023	Sur X1024	IC Mix X1025	Gases (Abs) X1026
VSTD005	5ul	0.5ul	0.25ul	0.5ul
10		1	0.5	1
20		2	1	2
30		3	1.5	3
40		4	2	4
50		5	2.5	5
60		6	3	6
80	⊥	8	4	8

Not calibrated for: Chloroethane, Acetone, Methylene Chloride

Changed quant ion for O-xylene (91 to 106) for Limu tissue analysis only. Standards were adjusted for 2g analysis of tissue samples and saved as SOIL_44HF

Continued on Page

Read and Understood By

US — Signed — 2/13/14 Date
GRC — Signed — 2/26/14 Date

PROJECT 2/13/14 AGI SOIL

Notebook No. 3317 55

Continued from Page

Run Method = SOIL_24.M
Tune = BFBSOIL.U EM = 1882
Quant Method = SOIL_43

5ul - ISTD = X1023
4ul - Sur = X1024

Injection Log

Directory: C:\MSDCHEM\1\DATA\021314

Line	Vial	FileName	Multiplier	SampleName	Misc Info
1	1	A0213301.d	0.	FLUSH	8260
2	1	A0213302.d	0.	50 NG BFB (2ul-X1007) - Pass 10:53am	8260
3	1	A0213303.d	0.	VSTD040/ICV C#5	8260
4	1	A0213304.d	0.	VSTD040/ICV - Pass	8260
5	1	A0213305.d	0.	VSTD000	8260
6	1	A0213306.d	0.	VSTD000	8260
7	1	A0213407.d	0.	VSTD000	8260
8	2	A0213408.d	0.	LOD 1	8260
9	3	A0213409.d	0.	LOD 1	8260
10	4	A0213410.d	0.	LOD 2	8260
11	5	A0213411.d	0.	LOD 2	8260
12	6	A0213412.d	0.	LOQ 1 > spiked wrong - RE	8260
13	7	A0213413.d	0.	LOQ 1	8260
14	8	A0213414.d	0.	VSTD040	8260

Conc. for SOIL_43
VSTD040
VSTD040

LS 2/26/14

LOD 1 = 5ul-X1023; 0.1ul-X1024; 20ul-X1029; 0.1ul-X1026

LOD 2 = 5ul-X1023; 0.2ul-X1024; 40ul-X1029; 0.2ul-X1026

LOQ 1 = 5ul-X1023; 0.5ul-X1024; 1ul-X1025; 0.5ul-X1026

ICV = 5ul-X1023; 4ul-X1024; 8ul-X1027; 4ul-X1028

LS 2/14/14

Continued on Page

Read and Understood By

Signed	Date	Signed	Date
LS	2/14/14	TRC	2/26/14

PROJECT 2/27/14 AGI 8260 SOIL

Notebook No. 3317 61

Continued from Page

Run Method = SOIL_24.m
Tune = BFBSOIL.U EM = 1887
Tekmar = NB3167 pg. 34
Quant Method = SOIL ~~43HI~~ 44HI LS 3/20/14

5ul ISTD = X1034
4ul Sur = X1035
15ul - Anti Foam ~~Agent~~ - X1040 LS 4/01/14

A0227401.D FLUSH
02.D FLUSH
03.D 50 NG BFB (2ul-X1007) Pass ~~8:43 am~~ 9:10 am LS 2/28/14
04.D VSTD040 100 LS 3/20/14 Pass
2ul - X1038 ; 4 ul X1039
05.D LCS Pass
8ul - X1036 ; 4 ul - X1037
06.D VSTD000 Pass
07.D 006613-01 MAK101L (1.98g) (A) I2↓ - RE 1g
08.D 006613-02 MAK102L (1.98g) I2↓ - RE 1g
09.D FLUSH
10.D 006613-01 MAK101L (1.02g) (A) I2↓ - RE
11.D 006613-02 MAK102L (1.01g) I2↓ - RE
12.D 006613-03 MAK105L (1.03g) I2↓ - RE
13.D 006613-04 MAK106L (0.99g) Pass
14.D 006613-05 MAK108, 109, 110, 113L (1.00g) Pass
15.D 006613-06 MAK111L (1.00g) Pass
16.D 006613-07 MAK112L (1.03g) I2↓ - RE
17.D 006613-08 MAK115L (0.98g) I2↓ - RE
18.D 006613-01MS MAK101LMS (0.99g) (B) I2↓ - RE
same as 05.D
19.D 006613-01MD MAK101MSD (1.00g) (C) I2↓ - RE
Same as 05.D
20.D VSTD040 100 LS 3/20/14
Same as 04.D
21.D VSTD040 100 LS 3/20/14
Same as 04.D

LS 2/28/14

Continued on Page

Read and Understood By

Signed	Date	Signed	Date
LS	2/28/14	DRC	3/4/14

Blank / Field Sample / QC Data

Sample Results

Chromatogram

Quant Report (If applicable)

Confirmation Analysis Quant Report / Chromatogram (If applicable)

Tentatively Identified Compounds (If required)

Rev. 1.1

ARDL, INC.
400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	NA	ARDL Lab No.:	006613-04K1
Desc/Location:	NA	Lab Filename:	A0227405
Sample Date:	NA	Received Date:	NA
Sample Time:	NA	Prep. Date:	02/27/2014
Matrix:	QC Material	Analysis Date:	02/27/2014
Amount Used:	2 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSW
% Moisture:	NA	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units
Toluene	2.50	12.5	96.1		UG/KG
Ethyl Benzene	2.50	12.5	97.1		UG/KG
m & p-Xylene	5.00	25.0	197		UG/KG
Styrene	2.50	12.5	96.2		UG/KG
o-Xylene	2.50	12.5	94.2		UG/KG
1,2,4-Trimethylbenzene	2.50	12.5	102		UG/KG

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	100%
4-Bromofluorobenzene	50-150	98%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Spike Blank for Run 0305JFSW, VOLATILES, GC/MS (8260) Page 1 of 1

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022714\
Data File : A0227405.D
Acq On : 27 Feb 2014 10:17 am
Operator :
Sample : ~~LGS~~ QD6613-04K1
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1 03/20/14

Quant Time: Mar 20 10:27:40 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:27:26 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Chlorobenzene-d5	8.77	82	251112	125.00 ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	270267	125.00 ug/Kg	0.00
System Monitoring Compounds					
2) Toluene-d8	7.83	98	454896	99.95 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 99.95%	
9) 4-Bromofluorobenzene	9.48	95	201347	97.89 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 97.89%	
Target Compounds					Qvalue
3) Toluene	7.86	91	584858	96.12 ug/Kg	99
4) Ethylbenzene	8.78	91	712422	97.06 ug/Kg	96
5) m & p-Xylene	8.86	91	1171976	197.09 ug/Kg	92
6) Styrene	9.15	104	428671	96.24 ug/Kg	89
7) o-Xylene	9.12	106	241786	94.24 ug/Kg	83
10) 1,2,4-Trimethylbenzene	9.88	105	741164	102.29 ug/Kg	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022714\
Data File : A0227405.D
Acq On : 27 Feb 2014 10:17 am
Operator :
Sample : ~~LCS~~ 006613-04K1
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1 JAS 03/20/14

Quant Time: Mar 20 10:27:40 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:27:26 2014
Response via : Initial Calibration



METHOD BLANK REPORT
ARDL, Inc. 400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	NA	ARDL Lab No.:	006613-04B1
Desc/Location:	NA	Lab Filename:	A0227406
Sample Date:	NA	Received Date:	NA
Sample Time:	NA	Prep. Date:	02/27/2014
Matrix:	QC Material	Analysis Date:	02/27/2014
Amount Used:	2 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSW
% Moisture:	NA	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units
Toluene	2.50	12.5	ND		UG/KG
Ethyl Benzene	2.50	12.5	ND		UG/KG
m & p-Xylene	5.00	25.0	ND		UG/KG
Styrene	2.50	12.5	ND		UG/KG
o-Xylene	2.50	12.5	ND		UG/KG
1,2,4-Trimethylbenzene	2.50	12.5	ND		UG/KG

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	101%
4-Bromofluorobenzene	50-150	111%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Blank for Run 0305JFSW, VOLATILES, GC/MS (8260) Page 1 of 1

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022714\
Data File : A0227406.D
Acq On : 27 Feb 2014 10:53 am
Operator :
Sample : ~~VSTD000~~ 006613-04B1
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1 JJS 03/20/14

Quant Time: Mar 20 12:52:58 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:27:26 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Chlorobenzene-d5	8.77	82	226740	125.00	ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	245868	125.00	ug/Kg	0.00
System Monitoring Compounds						
2) Toluene-d8	7.83	98	416493	101.35	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	101.35%	
9) 4-Bromofluorobenzene	9.48	95	208339	111.34	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	111.34%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022714\
Data File : A0227406.D
Acq On : 27 Feb 2014 10:53 am
Operator :
Sample : ~~VSTD000~~ 006613-04B1
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1 JB 03/20/14

Quant Time: Mar 20 12:52:58 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:27:26 2014
Response via : Initial Calibration



ARDL, INC.
400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	MAK106L	ARDL Lab No.:	006613-04
Desc/Location:	MAKAU MILITARY RESERV OAHU	Lab Filename:	A0227413
Sample Date:	01/14/2014	Received Date:	02/20/2014
Sample Time:	0845	Prep. Date:	02/27/2014
Matrix:	TISSUE	Analysis Date:	02/27/2014
Amount Used:	1 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSW
% Moisture:	No Moisture Present	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units	Dilution Factor
Toluene	5.0	25.0	ND		UG/KG	1
Ethyl Benzene	5.0	25.0	ND		UG/KG	1
m & p-Xylene	10.0	50.0	ND		UG/KG	1
Styrene	5.0	25.0	ND		UG/KG	1
o-Xylene	5.0	25.0	ND		UG/KG	1
1,2,4-Trimethylbenzene	5.0	25.0	ND		UG/KG	1

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	97%
4-Bromofluorobenzene	50-150	133%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Sample 006613-04, VOLATILES, GC/MS (8260) Page 1 of 1

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022714\
Data File : A0227413.D
Acq On : 27 Feb 2014 2:29 pm
Operator :
Sample : 006613-04
Misc : 8260 SOIL MAK106L (0.99g)
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Mar 20 12:53:53 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:27:26 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Chlorobenzene-d5	8.77	82	198739	125.00	ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	138647	125.00	ug/Kg	0.00
System Monitoring Compounds						
2) Toluene-d8	7.83	98	349119	96.92	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	96.92%	
9) 4-Bromofluorobenzene	9.48	95	139815	132.51	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	132.51%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022714\
Data File : A0227413.D
Acq On : 27 Feb 2014 2:29 pm
Operator :
Sample : 006613-04
Misc : 8260 SOIL MAK106L (0.99g)
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Mar 20 12:53:53 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:27:26 2014
Response via : Initial Calibration



ARDL, INC.
400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	MAK108,109,110,113L	ARDL Lab No.:	006613-05
Desc/Location:	MAKAU MILITARY RESERV OAHU	Lab Filename:	A0227414
Sample Date:	01/14/2014	Received Date:	02/20/2014
Sample Time:	---	Prep. Date:	02/27/2014
Matrix:	TISSUE	Analysis Date:	02/27/2014
Amount Used:	1 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSW
% Moisture:	No Moisture Present	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units	Dilution Factor
Toluene	5.0	25.0	ND		UG/KG	1
Ethyl Benzene	5.0	25.0	ND		UG/KG	1
m & p-Xylene	10.0	50.0	ND		UG/KG	1
Styrene	5.0	25.0	ND		UG/KG	1
o-Xylene	5.0	25.0	ND		UG/KG	1
1,2,4-Trimethylbenzene	5.0	25.0	ND		UG/KG	1

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	98%
4-Bromofluorobenzene	50-150	125%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Sample 006613-05, VOLATILES, GC/MS (8260) Page 1 of 1

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022714\
Data File : A0227414.D
Acq On : 27 Feb 2014 2:57 pm
Operator :
Sample : 006613-05
Misc : 8260 SOIL MAK108,109,110,113L (1.00g)
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Mar 20 12:54:33 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:27:26 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Chlorobenzene-d5	8.76	82	203750	125.00	ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	151097	125.00	ug/Kg	0.00
System Monitoring Compounds						
2) Toluene-d8	7.83	98	361246	97.82	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	97.82%	
9) 4-Bromofluorobenzene	9.48	95	144011	125.24	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	125.24%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022714\
Data File : A0227414.D
Acq On : 27 Feb 2014 2:57 pm
Operator :
Sample : 006613-05
Misc : 8260 SOIL MAK108,109,110,113L (1.00g)
ALS Vial : 8 Sample Multiplier: 1

Quant Time: Mar 20 12:54:33 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:27:26 2014
Response via : Initial Calibration



ARDL, INC.
400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	MAK111L	ARDL Lab No.:	006613-06
Desc/Location:	MAKAU MILITARY RESERV OAHU	Lab Filename:	A0227415
Sample Date:	01/14/2014	Received Date:	02/20/2014
Sample Time:	1205	Prep. Date:	02/27/2014
Matrix:	TISSUE	Analysis Date:	02/27/2014
Amount Used:	1 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSW
% Moisture:	No Moisture Present	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units	Dilution Factor
Toluene	5.0	25.0	ND		UG/KG	1
Ethyl Benzene	5.0	25.0	ND		UG/KG	1
m & p-Xylene	10.0	50.0	ND		UG/KG	1
Styrene	5.0	25.0	ND		UG/KG	1
o-Xylene	5.0	25.0	ND		UG/KG	1
1,2,4-Trimethylbenzene	5.0	25.0	ND		UG/KG	1

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	102%
4-Bromofluorobenzene	50-150	133%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Sample 006613-06, VOLATILES, GC/MS (8260) Page 1 of 1

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022714\
Data File : A0227415.D
Acq On : 27 Feb 2014 3:25 pm
Operator :
Sample : 006613-06
Misc : 8260 SOIL MAK111L (1.00g)
ALS Vial : 9 Sample Multiplier: 1

Quant Time: Mar 20 12:55:14 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:27:26 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Chlorobenzene-d5	8.76	82	197159	125.00	ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	144682	125.00	ug/Kg	0.00
System Monitoring Compounds						
2) Toluene-d8	7.83	98	365377	102.25	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	102.25%	
9) 4-Bromofluorobenzene	9.47	95	146367	132.93	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	132.93%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022714\
Data File : A0227415.D
Acq On : 27 Feb 2014 3:25 pm
Operator :
Sample : 006613-06
Misc : 8260 SOIL MAK111L (1.00g)
ALS Vial : 9 Sample Multiplier: 1

Quant Time: Mar 20 12:55:14 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:27:26 2014
Response via : Initial Calibration



Analytical Run Number 2

Calibration Information

Tune Data (If applicable)

Initial Calibration (If different than Initial Calibration in Run 1)
- Calibration Factor Summary
- Chromatogram
- Quant Report (If applicable)

Continuing Calibration
- Calibration Response Check
- Chromatogram
- Quant Report (If applicable)

Instrument Logs/Analysis Notebook Pages

Blank / Field Sample / QC Data

Sample Results
Chromatogram
Quant Report (If applicable)
Confirmation Analysis Quant Report / Chromatogram (If applicable)
Tentatively Identified Compounds (If required)

Rev. 1.1

Calibration Information

Tune Data (If applicable)

Initial Calibration (If different than Initial Calibration in Run 1)
- Calibration Factor Summary
- Chromatogram
- Quant Report (If applicable)

Continuing Calibration
- Calibration Response Check
- Chromatogram
- Quant Report (If applicable)

Rev. 1.1

5A

VOLATILE ORGANIC GC/MS TUNING AND MASS CALIBRATION - BROMOFLUOROBENZENE (BFB)

Lab Name: ARDL, Inc. Contract: ---

Lab Code: --- Case No.: 006613 SAS No.: --- SDG No.: ---

Lab File ID: A0228403 BFB Injection Date: 02/28/14

Instrument ID: AG1 BFB Injection Time: 10:13

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE	
50	8.0 - 40.0% of mass 95	16.4	
75	30.0 - 66.0% of mass 95	47.4	
95	Base Peak, 100% relative abundance	100.	
96	5.0 - 9.0% of mass 95	6.7	
173	Less than 2.0% of mass 174	0.0	(1)
174	Greater than 50.0% of mass 95	66.8	
175	4.0 - 9.0% of mass 174	6.9	(1)
176	Greater than 93.0%, but less that 101.0% of mass 174	97.3	(1)
177	5.0 - 9.0% of mass 176	6.4	(2)

(1)-Value is % mass 174 2-Value is % mass 176

THIS TUNE APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
1	VOA STD	100 STD	A0228404	02/28/14	10:36
2	SPBK 02/28/14	006613-01K1	A0228405	02/28/14	11:16
3	BLANK 02/28/14	006613-01B1	A0228406	02/28/14	11:51
4	MAK101L	006613-01	A0228407	02/28/14	12:25
5	MAK102L	006613-02	A0228408	02/28/14	13:03
6	MAK105L	006613-03	A0228409	02/28/14	13:44
7	MAK112L	006613-07	A0228410	02/28/14	14:22
8	MAK115L	006613-08	A0228411	02/28/14	14:50
9					
10					
11					
12					
13					
14					
15					
16					
17					
18					
19					
20					
21					
22					

Data Path : C:\MSDCHEM\1\DATA\022814\
Data File : A0228403.D
Acq On : 28 Feb 2014 10:13 am
Operator :
Sample : 50 NG BFB
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1

Integration File: rteint.p

Method : C:\MSDCHEM\1\METHODS\BFBSOIL.M
Title :
Last Update :



AutoFind: Scans 499, 500, 501; Background Corrected with Scan 491

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	16.4	7207	PASS
75	95	30	66	47.4	20786	PASS
95	95	100	100	100.0	43858	PASS
96	95	5	9	6.7	2953	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	66.8	29298	PASS
175	174	4	9	6.9	2009	PASS
176	174	93	101	97.3	28509	PASS
177	176	5	9	6.4	1811	PASS

Average of 4.683 to 4.690 min.: A0228403.D
50 NG BFB
Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
36.10	181	62.10	1738	88.10	1296	177.00	1811
37.15	1501	63.15	1479	92.05	1220		
38.15	1343	68.10	3278	93.10	1749		
39.10	747	69.10	4532	94.15	4817		
47.10	174	73.10	1968	95.10	43858		
49.10	1875	74.10	7570	96.10	2953		
50.10	7207	75.10	20786	140.90	178		
51.10	2409	76.10	1804	143.00	192		
56.10	421	79.00	453	174.00	29298		
57.10	840	81.00	819	175.05	2009		
61.10	1869	87.05	1667	176.00	28509		

Continuing Calibration Report AG1

Method Path : C:\MSDCHEM\1\METHODS\
Method File : SOIL_44HI.M
Title : 8260 VOLATILES
Last Update : Thu Mar 20 10:38:54 2014
Response Via : Initial Calibration

CC Data File: A0228404.D

Min. RRF : 0.000 Min. Rel. Area : 50%
Max. RRF Dev : 20% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%
1	I	Chlorobenzene-d5	1.000	1.000	0.0	80
2	S	Toluene-d8	2.266	2.276	-0.5	78
3	MCT	Toluene	3.029	3.058	-1.0	83
4	CT	Ethylbenzene	3.654	3.655	-0.0	84
5	T	m & p-Xylene	2.960	2.908	1.7	82
6	T	Styrene	2.217	2.184	1.5	80
7	T	o-Xylene	1.277	1.267	0.8	81
8	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	69
9	S	4-Bromofluorobenzene	0.951	1.063	-11.7	75
10	T	1,2,4-Trimethylbenzene	3.351	3.813	-13.8	79

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022814\
Data File : A0228404.D
Acq On : 28 Feb 2014 10:36 am
Operator :
Sample : VSTD100
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 20 10:38:50 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:36:43 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Chlorobenzene-d5	8.76	82	224170	125.00	ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	184479	125.00	ug/Kg	0.00
System Monitoring Compounds						
2) Toluene-d8	7.83	98	408180	100.46	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	100.46%	
9) 4-Bromofluorobenzene	9.48	95	156826	111.70	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	111.70%	
Target Compounds						Qvalue
3) Toluene	7.86	91	548484	100.98	ug/Kg	100
4) Ethylbenzene	8.77	91	655494	100.04	ug/Kg	95
5) m & p-Xylene	8.86	91	1043122	196.50	ug/Kg	92
6) Styrene	9.14	104	391697	98.50	ug/Kg	92
7) o-Xylene	9.11	106	227193	99.20	ug/Kg	81
10) 1,2,4-Trimethylbenzene	9.88	105	562808	113.79	ug/Kg	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022814\
Data File : A0228404.D
Acq On : 28 Feb 2014 10:36 am
Operator :
Sample : VSTD100
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 20 10:38:50 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:36:43 2014
Response via : Initial Calibration



Instrument Logs/
Analysis Notebook Pages

Rev. 1.1

62
PROJECT 2/28/14 AGI 8260 SOIL

Notebook No. 3317
Continued from Page

Run Method = SOIL_24.M
Tune = BFBSOIL.U EM = 1887
Telmar = NB 3167 pg. 34
Quant Method = SOIL_~~43HI~~ 44HI
LS 3/20/14

5ul - ISTD = X1034
4ul - Sur = X1035
15ul - Anti Foam Agent - X1040

A072840 01.D FLUSH
02.D 50 NG BFB (2ul - X1007) Fails
03.D 50 NG BFB 1 Pass 10:13am
04.D VSTD~~040~~100 LS 3/20/14 Pass
2ul - X1038
05.D LCS Pass
8 ul - X1036
06.D VSTD000 Pass
07.D 006613-01 MAK101L (0.98g) (A) Pass
08.D 006613-02 MAK102L (1.00g) Pass
09.D 006613-03 MAK105L (0.98g) Pass
10.D 006613-07 MAK112L (0.98g) Pass
11.D 006613-08 MAK115L (0.99g) Pass
12.D 006613-01MS MAK101LMS (1.00g) (B) MPV - RE
Same as 05.D
13.D 006613-01MD MAK101LMSD (0.98g) (C) C#445 ↓ - RE
Same as 05.D
14.D VSTD~~040~~100 LS 3/20/14
Same as 04.D
15.D VSTD~~040~~100 LS 3/20/14
Same as 04.D

LS 3/03/14

Continued on Page

Read and Understood By

Signed	Date	Signed	Date
LS	3/03/14	JRC	3/4/14

Blank / Field Sample / QC Data

Sample Results

Chromatogram

Quant Report (If applicable)

Confirmation Analysis Quant Report / Chromatogram (If applicable)

Tentatively Identified Compounds (If required)

Rev. 1.1

ARDL, INC.
400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	NA	ARDL Lab No.:	006613-01K1
Desc/Location:	NA	Lab Filename:	A0228405
Sample Date:	NA	Received Date:	NA
Sample Time:	NA	Prep. Date:	02/28/2014
Matrix:	QC Material	Analysis Date:	02/28/2014
Amount Used:	2 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSX
% Moisture:	NA	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units
Toluene	2.50	12.5	95.0		UG/KG
Ethyl Benzene	2.50	12.5	93.5		UG/KG
m & p-Xylene	5.00	25.0	185		UG/KG
Styrene	2.50	12.5	93.9		UG/KG
o-Xylene	2.50	12.5	92.8		UG/KG
1,2,4-Trimethylbenzene	2.50	12.5	111		UG/KG

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	102%
4-Bromofluorobenzene	50-150	114%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Spike Blank for Run 0305JFSX, VOLATILES, GC/MS (8260) Page 1 of 1

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022814\
Data File : A0228405.D
Acq On : 28 Feb 2014 11:16 am
Operator :
Sample : ~~Les~~ 006613-01K1
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1 03/20/14

Quant Time: Mar 20 10:39:02 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:38:54 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Chlorobenzene-d5	8.76	82	241311	125.00	ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	207659	125.00	ug/Kg	0.00
System Monitoring Compounds						
2) Toluene-d8	7.83	98	444750	101.69	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	101.69%	
9) 4-Bromofluorobenzene	9.47	95	180000	113.90	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	113.90%	
Target Compounds						Qvalue
3) Toluene	7.86	91	555290	94.97	ug/Kg	99
4) Ethylbenzene	8.77	91	659279	93.47	ug/Kg	96
5) m & p-Xylene	8.85	91	1056671	184.92	ug/Kg	93
6) Styrene	9.14	104	401802	93.87	ug/Kg	92
7) o-Xylene	9.11	106	228736	92.78	ug/Kg	81
10) 1,2,4-Trimethylbenzene	9.87	105	618817	111.15	ug/Kg	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022814\
Data File : A0228405.D
Acq On : 28 Feb 2014 11:16 am
Operator :
Sample : ~~Les~~ 006613-01K1
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1 03/20/14

Quant Time: Mar 20 10:39:02 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:38:54 2014
Response via : Initial Calibration



METHOD BLANK REPORT
ARDL, Inc. 400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	NA	ARDL Lab No.:	006613-01B1
Desc/Location:	NA	Lab Filename:	A0228406
Sample Date:	NA	Received Date:	NA
Sample Time:	NA	Prep. Date:	02/28/2014
Matrix:	QC Material	Analysis Date:	02/28/2014
Amount Used:	2 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSX
% Moisture:	NA	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units
Toluene	2.50	12.5	ND		UG/KG
Ethyl Benzene	2.50	12.5	ND		UG/KG
m & p-Xylene	5.00	25.0	ND		UG/KG
Styrene	2.50	12.5	ND		UG/KG
o-Xylene	2.50	12.5	ND		UG/KG
1,2,4-Trimethylbenzene	2.50	12.5	ND		UG/KG

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	99%
4-Bromofluorobenzene	50-150	116%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Blank for Run 0305JFSX, VOLATILES, GC/MS (8260) Page 1 of 1

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022814\
Data File : A0228406.D
Acq On : 28 Feb 2014 11:51 am
Operator :
Sample : V~~STD000~~ 006613-01B1
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1 03/20/14

Quant Time: Mar 20 11:44:40 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:38:54 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Chlorobenzene-d5	8.76	82	244246	125.00	ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	233307	125.00	ug/Kg	0.00
System Monitoring Compounds						
2) Toluene-d8	7.83	98	436952	98.70	ug/Kg	0.00
Spiked Amount 100.000			Recovery = 98.70%			
9) 4-Bromofluorobenzene	9.47	95	205141	115.54	ug/Kg	0.00
Spiked Amount 100.000			Recovery = 115.54%			

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022814\
Data File : A0228406.D
Acq On : 28 Feb 2014 11:51 am
Operator :
Sample : ~~VSTD000~~ D06613-01B1
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1 JB 03/20/14

Quant Time: Mar 20 11:44:40 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:38:54 2014
Response via : Initial Calibration



ARDL, INC.
400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	MAK101L	ARDL Lab No.:	006613-01
Desc/Location:	MAKAU MILITARY RESERV OAHU	Lab Filename:	A0228407
Sample Date:	01/13/2014	Received Date:	02/20/2014
Sample Time:	0910	Prep. Date:	02/28/2014
Matrix:	TISSUE	Analysis Date:	02/28/2014
Amount Used:	1 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSX
% Moisture:	No Moisture Present	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units	Dilution Factor
Toluene	5.0	25.0	ND		UG/KG	1
Ethyl Benzene	5.0	25.0	ND		UG/KG	1
m & p-Xylene	10.0	50.0	ND		UG/KG	1
Styrene	5.0	25.0	ND		UG/KG	1
o-Xylene	5.0	25.0	ND		UG/KG	1
1,2,4-Trimethylbenzene	5.0	25.0	ND		UG/KG	1

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	101%
4-Bromofluorobenzene	50-150	131%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022814\
Data File : A0228407.D
Acq On : 28 Feb 2014 12:25 pm
Operator :
Sample : 006613-01
Misc : 8260 SOIL MAK101L (0.98g)
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 20 10:41:18 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:38:54 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Chlorobenzene-d5	8.77	82	234011	125.00	ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	145054	125.00	ug/Kg	0.00
System Monitoring Compounds						
2) Toluene-d8	7.83	98	426718	100.61	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	100.61%	
9) 4-Bromofluorobenzene	9.48	95	144218	130.64	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	130.64%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022814\
Data File : A0228407.D
Acq On : 28 Feb 2014 12:25 pm
Operator :
Sample : 006613-01
Misc : 8260 SOIL MAK101L (0.98g)
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 20 10:41:18 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:38:54 2014
Response via : Initial Calibration



ARDL, INC.
400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	MAK102L	ARDL Lab No.:	006613-02
Desc/Location:	MAKAU MILITARY RESERV OAHU	Lab Filename:	A0228408
Sample Date:	01/13/2014	Received Date:	02/20/2014
Sample Time:	1002	Prep. Date:	02/28/2014
Matrix:	TISSUE	Analysis Date:	02/28/2014
Amount Used:	1 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSX
% Moisture:	No Moisture Present	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units	Dilution Factor
Toluene	5.0	25.0	ND		UG/KG	1
Ethyl Benzene	5.0	25.0	ND		UG/KG	1
m & p-Xylene	10.0	50.0	ND		UG/KG	1
Styrene	5.0	25.0	ND		UG/KG	1
o-Xylene	5.0	25.0	ND		UG/KG	1
1,2,4-Trimethylbenzene	5.0	25.0	ND		UG/KG	1

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	98%
4-Bromofluorobenzene	50-150	116%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Sample 006613-02, VOLATILES, GC/MS (8260) Page 1 of 1

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022814\
Data File : A0228408.D
Acq On : 28 Feb 2014 1:03 pm
Operator :
Sample : 006613-02
Misc : 8260 SOIL MAK102L (1.00g)
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 20 10:41:26 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:38:54 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Chlorobenzene-d5	8.76	82	255914	125.00	ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	191516	125.00	ug/Kg	0.00
System Monitoring Compounds						
2) Toluene-d8	7.83	98	454976	98.09	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	98.09%	
9) 4-Bromofluorobenzene	9.48	95	168745	115.78	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	115.78%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022814\
Data File : A0228408.D
Acq On : 28 Feb 2014 1:03 pm
Operator :
Sample : 006613-02
Misc : 8260 SOIL MAK102L (1.00g)
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 20 10:41:26 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:38:54 2014
Response via : Initial Calibration



ARDL, INC.
400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	MAK105L	ARDL Lab No.:	006613-03
Desc/Location:	MAKAU MILITARY RESERV OAHU	Lab Filename:	A0228409
Sample Date:	01/13/2014	Received Date:	02/20/2014
Sample Time:	1520	Prep. Date:	02/28/2014
Matrix:	TISSUE	Analysis Date:	02/28/2014
Amount Used:	1 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSX
% Moisture:	No Moisture Present	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units	Dilution Factor
Toluene	5.0	25.0	ND		UG/KG	1
Ethyl Benzene	5.0	25.0	ND		UG/KG	1
m & p-Xylene	10.0	50.0	ND		UG/KG	1
Styrene	5.0	25.0	ND		UG/KG	1
o-Xylene	5.0	25.0	ND		UG/KG	1
1,2,4-Trimethylbenzene	5.0	25.0	ND		UG/KG	1

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	97%
4-Bromofluorobenzene	50-150	119%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022814\
Data File : A0228409.D
Acq On : 28 Feb 2014 1:44 pm
Operator :
Sample : 006613-03
Misc : 8260 SOIL MAK105L (0.98g)
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 20 11:45:28 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:38:54 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Chlorobenzene-d5	8.76	82	248663	125.00	ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	177554	125.00	ug/Kg	0.00
System Monitoring Compounds						
2) Toluene-d8	7.83	98	438876	97.38	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	97.38%	
9) 4-Bromofluorobenzene	9.48	95	160488	118.77	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	118.77%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022814\
Data File : A0228409.D
Acq On : 28 Feb 2014 1:44 pm
Operator :
Sample : 006613-03
Misc : 8260 SOIL MAK105L (0.98g)
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 20 11:45:28 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:38:54 2014
Response via : Initial Calibration



ARDL, INC.
400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	MAK112L	ARDL Lab No.:	006613-07
Desc/Location:	MAKAU MILITARY RESERV OAHU	Lab Filename:	A0228410
Sample Date:	01/14/2014	Received Date:	02/20/2014
Sample Time:	1250	Prep. Date:	02/28/2014
Matrix:	TISSUE	Analysis Date:	02/28/2014
Amount Used:	1 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSX
% Moisture:	No Moisture Present	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units	Dilution Factor
Toluene	5.0	25.0	ND		UG/KG	1
Ethyl Benzene	5.0	25.0	ND		UG/KG	1
m & p-Xylene	10.0	50.0	ND		UG/KG	1
Styrene	5.0	25.0	ND		UG/KG	1
o-Xylene	5.0	25.0	ND		UG/KG	1
1,2,4-Trimethylbenzene	5.0	25.0	ND		UG/KG	1

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	98%
4-Bromofluorobenzene	50-150	129%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Sample 006613-07, VOLATILES, GC/MS (8260) Page 1 of 1

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022814\
Data File : A0228410.D
Acq On : 28 Feb 2014 2:22 pm
Operator :
Sample : 006613-07
Misc : 8260 SOIL MAK112L (0.98g)
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 20 11:46:04 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:38:54 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Chlorobenzene-d5	8.76	82	226731	125.00 ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	157537	125.00 ug/Kg	0.00
System Monitoring Compounds					
2) Toluene-d8	7.83	98	400796	97.53 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 97.53%	
9) 4-Bromofluorobenzene	9.48	95	154556	128.91 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 128.91%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022814\
Data File : A0228410.D
Acq On : 28 Feb 2014 2:22 pm
Operator :
Sample : 006613-07
Misc : 8260 SOIL MAK112L (0.98g)
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 20 11:46:04 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:38:54 2014
Response via : Initial Calibration



ARDL, INC.
400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/20/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	MAK115L	ARDL Lab No.:	006613-08
Desc/Location:	MAKAU MILITARY RESERV OAHU	Lab Filename:	A0228411
Sample Date:	01/14/2014	Received Date:	02/20/2014
Sample Time:	1400	Prep. Date:	02/28/2014
Matrix:	TISSUE	Analysis Date:	02/28/2014
Amount Used:	1 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSX
% Moisture:	No Moisture Present	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units	Dilution Factor
Toluene	5.0	25.0	ND		UG/KG	1
Ethyl Benzene	5.0	25.0	ND		UG/KG	1
m & p-Xylene	10.0	50.0	ND		UG/KG	1
Styrene	5.0	25.0	ND		UG/KG	1
o-Xylene	5.0	25.0	ND		UG/KG	1
1,2,4-Trimethylbenzene	5.0	25.0	ND		UG/KG	1

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	99%
4-Bromofluorobenzene	50-150	124%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Sample 006613-08, VOLATILES, GC/MS (8260) Page 1 of 1

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022814\
Data File : A0228411.D
Acq On : 28 Feb 2014 2:50 pm
Operator :
Sample : 006613-08
Misc : 8260 SOIL MAK115L (0.99g)
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Mar 20 10:41:59 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:38:54 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Chlorobenzene-d5	8.76	82	252826	125.00 ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	185155	125.00 ug/Kg	0.00
System Monitoring Compounds					
2) Toluene-d8	7.83	98	453126	98.88 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 98.88%	
9) 4-Bromofluorobenzene	9.48	95	174471	123.82 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 123.82%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\022814\
Data File : A0228411.D
Acq On : 28 Feb 2014 2:50 pm
Operator :
Sample : 006613-08
Misc : 8260 SOIL MAK115L (0.99g)
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Mar 20 10:41:59 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 10:38:54 2014
Response via : Initial Calibration



Analytical Run Number 3

Calibration Information

Tune Data (If applicable)

Initial Calibration (If different than Initial Calibration in Run 1)
- Calibration Factor Summary
- Chromatogram
- Quant Report (If applicable)

Continuing Calibration
- Calibration Response Check
- Chromatogram
- Quant Report (If applicable)

Instrument Logs/Analysis Notebook Pages

Blank / Field Sample / QC Data

Sample Results
Chromatogram
Quant Report (If applicable)
Confirmation Analysis Quant Report / Chromatogram (If applicable)
Tentatively Identified Compounds (If required)

Rev. 1.1

Calibration Information

Tune Data (If applicable)

Initial Calibration (If different than Initial Calibration in Run 1)
- Calibration Factor Summary
- Chromatogram
- Quant Report (If applicable)

Continuing Calibration
- Calibration Response Check
- Chromatogram
- Quant Report (If applicable)

Rev. 1.1

5A

VOLATILE ORGANIC GC/MS TUNING AND MASS CALIBRATION - BROMOFLUOROBENZENE (BFB)

Lab Name: ARDL, Inc. Contract: ---

Lab Code: --- Case No.: 006613 SAS No.: --- SDG No.: ---

Lab File ID: A0303403 BFB Injection Date: 03/03/14

Instrument ID: AG1 BFB Injection Time: 10:37

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE	
50	8.0 - 40.0% of mass 95	16.4	
75	30.0 - 66.0% of mass 95	46.4	
95	Base Peak, 100% relative abundance	100.	
96	5.0 - 9.0% of mass 95	6.8	
173	Less than 2.0% of mass 174	0.0	(1)
174	Greater than 50.0% of mass 95	69.6	
175	4.0 - 9.0% of mass 174	7.0	(1)
176	Greater than 93.0%, but less that 101.0% of mass 174	96.2	(1)
177	5.0 - 9.0% of mass 176	6.4	(2)

(1)-Value is % mass 174 2-Value is % mass 176

THIS TUNE APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
1	VOA STD	100 STD	A0303404	03/03/14	11:02
2	SPBK 03/03/14	006613-01K2	A0303405	03/03/14	11:38
3	BLANK 03/03/14	006613-01B2	A0303406	03/03/14	12:15
4	MAK101LMS	006613-01MS	A0303407	03/03/14	12:50
5	MAK101LMSD	006613-01MD	A0303408	03/03/14	13:26
6					
7					
8					
9					
10					
11					
12					
13					
14					
15					
16					
17					
18					
19					
20					
21					
22					

Data Path : C:\MSDCHEM\1\DATA\030314\
Data File : A0303403.D
Acq On : 3 Mar 2014 10:37 am
Operator :
Sample : 50 NG BFB
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1

Integration File: rteint.p

Method : C:\MSDCHEM\1\METHODS\BFBSOIL.M
Title :
Last Update :





AutoFind: Scans 501, 502, 503; Background Corrected with Scan 492

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	16.4	10287	PASS
75	95	30	66	46.4	29128	PASS
95	95	100	100	100.0	62714	PASS
96	95	5	9	6.8	4256	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	69.6	43650	PASS
175	174	4	9	7.0	3049	PASS
176	174	93	101	96.2	42008	PASS
177	176	5	9	6.4	2708	PASS

Average of 4.689 to 4.696 min.: A0303403.D
50 NG BFB
Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
36.10	202	56.10	708	76.10	2389	95.10	62714
37.15	2604	57.10	1407	77.20	184	96.10	4256
38.15	1955	61.10	2230	79.05	1103	143.00	169
39.10	958	62.10	2613	80.00	209	174.00	43650
40.00	52	63.10	1863	81.00	1303	175.00	3049
44.05	362	68.10	5123	81.90	173	176.00	42008
47.05	686	69.10	6027	87.05	2271	176.95	2708
48.10	171	70.10	407	88.05	2061		
49.10	1896	73.10	2635	92.00	1474		
50.10	10287	74.10	9059	93.05	2971		
51.10	3158	75.10	29128	94.10	6986		

Continuing Calibration Report AG1

Method Path : C:\MSDCHEM\1\METHODS\
Method File : SOIL_44HI.M
Title : 8260 VOLATILES
Last Update : Thu Mar 20 11:40:32 2014
Response Via : Initial Calibration

CC Data File: A0303404.D

Min. RRF : 0.000 Min. Rel. Area : 50%
Max. RRF Dev : 20% Max. Rel. Area : 150%

		Compound	AvgRF	CCRF	%Dev	Area%
1	I	Chlorobenzene-d5	1.000	1.000	0.0	99
2	S	Toluene-d8	2.266	2.290	-1.1	96
3	MCT	Toluene	3.029	3.271	-8.0	110
4	CT	Ethylbenzene	3.654	3.833	-4.9	109
5	T	m & p-Xylene	2.960	3.008	-1.6	105
6	T	Styrene	2.217	2.308	-4.1	104
7	T	o-Xylene	1.277	1.369	-7.2	107
8	I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92
9	S	4-Bromofluorobenzene	0.951	1.028	-8.1	97
10	T	1,2,4-Trimethylbenzene	3.351	3.665	-9.4	101

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\030314\
Data File : A0303404.D
Acq On : 3 Mar 2014 11:02 am
Operator :
Sample : VSTD100
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 20 11:40:19 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 11:39:15 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Chlorobenzene-d5	8.77	82	276553	125.00 ug/Kg	0.01
8) 1,4-Dichlorobenzene-d4	10.15	152	247227	125.00 ug/Kg	0.00
System Monitoring Compounds					
2) Toluene-d8	7.84	98	506584	101.06 ug/Kg	0.01
Spiked Amount 100.000			Recovery	= 101.06%	
9) 4-Bromofluorobenzene	9.48	95	203385	108.10 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 108.10%	
Target Compounds					Qvalue
3) Toluene	7.87	91	723633	107.99 ug/Kg	100
4) Ethylbenzene	8.78	91	847987	104.90 ug/Kg	96
5) m & p-Xylene	8.87	91	1331028	203.24 ug/Kg	94
6) Styrene	9.15	104	510684	104.10 ug/Kg	95
7) o-Xylene	9.12	106	302912	107.21 ug/Kg	85
10) 1,2,4-Trimethylbenzene	9.88	105	724839	109.36 ug/Kg	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\030314\
Data File : A0303404.D
Acq On : 3 Mar 2014 11:02 am
Operator :
Sample : VSTD100
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 20 11:40:19 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 11:39:15 2014
Response via : Initial Calibration



Instrument Logs/
Analysis Notebook Pages

Rev. 1.1

PROJECT 3/03/14 AGI 8260 SOIL

Run Method = SOIL_24.M
Tune = BFBSOIL.U EM = 1887
Telkmar = NB 3167 pg. 34
Quant Method = SOIL_~~43HI~~ 44HI LS 3/20/14

5ul - ISTD = X1049
4ul - Sur = X1050
15ul - Anti Foam ~~Acetone~~ - X1040 LS 4/01/14

A0303401.D FLUSH
02.D 50 NG BFB (2ul - X1007) Fails
03.D 50 NG BFB Pass 10:37 am
04.D VSTD~~040~~100 LS 3/20/14 Pass
2ul - X1038
05.D LCS Pass
8ul - X1051
06.D VSTD000 Pass
07.D 006613-01MS MAK10ILMS (0.98g) (B) Report
Same as 04.D
08.D 006613-01MD MAK10ILMSD (0.99g) (C) Report
Same as 04.D
09.D VSTD~~040~~100 LS 3/20/14
Same as 04.D

LS 3/03/14

Continued on Page

Read and Understood By

LS Signed 3/03/14 Date ~~TRC~~ Signed 3/4/14 Date

Blank / Field Sample / QC Data

Sample Results

Chromatogram

Quant Report (If applicable)

Confirmation Analysis Quant Report / Chromatogram (If applicable)

Tentatively Identified Compounds (If required)

Rev. 1.1

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\030314\
Data File : A0303405.D
Acq On : 3 Mar 2014 11:38 am
Operator :
Sample : ~~LCS~~ 006613-01K2
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1 SAS 03/20/14

Quant Time: Mar 20 11:40:48 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 11:40:32 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Chlorobenzene-d5	8.77	82	270704	125.00	ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	257087	125.00	ug/Kg	0.00
System Monitoring Compounds						
2) Toluene-d8	7.84	98	562705	114.69	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	114.69%	
9) 4-Bromofluorobenzene	9.48	95	226485	115.76	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	115.76%	
Target Compounds						Qvalue
3) Toluene	7.87	91	693949	105.80	ug/Kg	99
4) Ethylbenzene	8.78	91	800556	101.17	ug/Kg	97
5) m & p-Xylene	8.86	91	1274482	198.81	ug/Kg	94
6) Styrene	9.15	104	489099	101.85	ug/Kg	95
7) o-Xylene	9.12	106	284214	102.76	ug/Kg	86
10) 1,2,4-Trimethylbenzene	9.88	105	767848	111.40	ug/Kg	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\030314\
Data File : A0303405.D
Acq On : 3 Mar 2014 11:38 am
Operator :
Sample : ~~LCS~~ QQ6613-01K2
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1 JJ 03/20/14

Quant Time: Mar 20 11:40:48 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 11:40:32 2014
Response via : Initial Calibration



Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\030314\
Data File : A0303406.D
Acq On : 3 Mar 2014 12:15 pm
Operator :
Sample : VS~~TB000~~ 006613-01B2
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1 03/20/14

Quant Time: Mar 20 11:42:30 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 11:40:32 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc Units	Dev(Min)
1) Chlorobenzene-d5	8.77	82	284953	125.00 ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.14	152	273520	125.00 ug/Kg	0.00
System Monitoring Compounds					
2) Toluene-d8	7.83	98	554325	107.33 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 107.33%	
9) 4-Bromofluorobenzene	9.48	95	230807	110.88 ug/Kg	0.00
Spiked Amount 100.000			Recovery	= 110.88%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (QT Reviewed)

Data Path : C:\MSDCHEM\1\DATA\030314\
Data File : A0303406.D
Acq On : 3 Mar 2014 12:15 pm
Operator :
Sample : ~~VSTD000~~ 006613-01B2
Misc : 8260 SOIL
ALS Vial : 1 Sample Multiplier: 1 03/20/14

Quant Time: Mar 20 11:42:30 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 11:40:32 2014
Response via : Initial Calibration



ARDL, INC.
400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/21/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	MAK101L	ARDL Lab No.:	006613-01MS
Desc/Location:	MAKAU MILITARY RESERV OAHU	Lab Filename:	A0303407
Sample Date:	01/13/2014	Received Date:	02/20/2014
Sample Time:	0910	Prep. Date:	03/03/2014
Matrix:	TISSUE	Analysis Date:	03/03/2014
Amount Used:	1 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSX
% Moisture:	No Moisture Present	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units	Dilution Factor
Toluene	5.00	25.0	183		UG/KG	1
Ethyl Benzene	5.00	25.0	173		UG/KG	1
m & p-Xylene	10.0	50.0	340		UG/KG	1
Styrene	5.00	25.0	40.1		UG/KG	1
o-Xylene	5.00	25.0	129		UG/KG	1
1,2,4-Trimethylbenzene	5.00	25.0	179		UG/KG	1

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	97%
4-Bromofluorobenzene	50-150	108%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Matrix Spike for 006613-01, VOLATILES, GC/MS (8260) Page 1 of 1

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\030314\
Data File : A0303407.D
Acq On : 3 Mar 2014 12:50 pm
Operator :
Sample : 006613-01MS
Misc : 8260 SOIL MAK101LMS (0.98g)
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 20 11:41:23 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 11:40:32 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Chlorobenzene-d5	8.77	82	349346	125.00	ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.15	152	252292	125.00	ug/Kg	0.00
System Monitoring Compounds						
2) Toluene-d8	7.83	98	611681	96.60	ug/Kg	0.00
Spiked Amount 100.000			Recovery = 96.60%			
9) 4-Bromofluorobenzene	9.48	95	208275	108.48	ug/Kg	0.00
Spiked Amount 100.000			Recovery = 108.48%			
Target Compounds						Qvalue
3) Toluene	7.86	91	775403	91.60	ug/Kg	100
4) Ethylbenzene	8.78	91	882255	86.40	ug/Kg	97
5) m & p-Xylene	8.86	91	1404456	169.77	ug/Kg	94
6) Styrene	9.17	104	124287	20.06	ug/Kg	96
7) o-Xylene	9.17	106	230094	64.47	ug/Kg#	1
10) 1,2,4-Trimethylbenzene	9.88	105	606139	89.61	ug/Kg	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\030314\
Data File : A0303407.D
Acq On : 3 Mar 2014 12:50 pm
Operator :
Sample : 006613-01MS
Misc : 8260 SOIL MAK101LMS (0.98g)
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 20 11:41:23 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Thu Mar 20 11:40:32 2014
Response via : Initial Calibration



ARDL, INC.
400 Aviation Drive; P.O. Box 1566
Mt. Vernon, Illinois 62864

Lab Report No: 006613 Report Date: 03/21/2014

Project Name: MAHAU MILITARY RESER Analysis: VOLATILES, GC/MS (8260)
Project No.: Analytical Method: 8260B
NELAC Certified - IL100308 Prep Method: 5030A

Field ID:	MAK101L	ARDL Lab No.:	006613-01MD
Desc/Location:	MAKAU MILITARY RESERV OAHU	Lab Filename:	A0303408
Sample Date:	01/13/2014	Received Date:	02/20/2014
Sample Time:	0910	Prep. Date:	03/03/2014
Matrix:	TISSUE	Analysis Date:	03/03/2014
Amount Used:	1 g	Instrument ID:	AG1
Final Volume:	5 mL	QC Batch:	0305JFSX
% Moisture:	No Moisture Present	Level:	LOW

Parameter	Method Limit	Reporting Limit	Result	Data Flag	Units	Dilution Factor
Toluene	5.00	25.0	189		UG/KG	1
Ethyl Benzene	5.00	25.0	174		UG/KG	1
m & p-Xylene	10.0	50.0	339		UG/KG	1
Styrene	5.00	25.0	36.5		UG/KG	1
o-Xylene	5.00	25.0	122		UG/KG	1
1,2,4-Trimethylbenzene	5.00	25.0	177		UG/KG	1

SURROGATE RECOVERIES:	Limits	Results
Toluene-d8	50-150	110%
4-Bromofluorobenzene	50-150	114%

Surrogate recoveries marked with '*' indicates they are outside standard limits.

Matrix Spike Duplicate for 006613-01, VOLATILES, GC/MS (8260) Page 1 of 1

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\030314\
Data File : A0303408.D
Acq On : 3 Mar 2014 1:26 pm
Operator :
Sample : 006613-01MD
Misc : 8260 SOIL MAK101LMSD (0.99g)
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 21 09:23:06 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Fri Mar 21 09:23:02 2014
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Chlorobenzene-d5	8.77	82	313113	125.00	ug/Kg	0.00
8) 1,4-Dichlorobenzene-d4	10.15	152	260288	125.00	ug/Kg	0.00
System Monitoring Compounds						
2) Toluene-d8	7.83	98	621890	109.58	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	109.58%	
9) 4-Bromofluorobenzene	9.48	95	225846	114.01	ug/Kg	0.00
Spiked Amount 100.000			Recovery	=	114.01%	
Target Compounds						Qvalue
3) Toluene	7.86	91	715339	94.29	ug/Kg	99
4) Ethylbenzene	8.78	91	794660	86.83	ug/Kg	97
5) m & p-Xylene	8.86	91	1257783	169.63	ug/Kg	94
6) Styrene	9.17	104	101295	18.24	ug/Kg	99
7) o-Xylene	9.17	106	194540	60.81	ug/Kg#	1
10) 1,2,4-Trimethylbenzene	9.88	105	618396	88.62	ug/Kg	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quantitation Report (Not Reviewed)

Data Path : C:\MSDCHEM\1\DATA\030314\
Data File : A0303408.D
Acq On : 3 Mar 2014 1:26 pm
Operator :
Sample : 006613-01MD
Misc : 8260 SOIL MAK101LMSD (0.99g)
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 21 09:23:06 2014
Quant Method : C:\MSDCHEM\1\METHODS\SOIL_44HI.M
Quant Title : 8260 VOLATILES
QLast Update : Fri Mar 21 09:23:02 2014
Response via : Initial Calibration



Example Calculation

Rev. 1.1

EXAMPLE CALCULATIONS

Response Factor Calculations

The data system tabulates the area response of the characteristic ions against the concentration for each compound and each internal standard. Response factors (RF) are calculated for each compound relative to the internal standard assignments. The RF is calculated as follows:

$$RF = (A_X C_{is})/A_{is} C_X)$$

where:

A_X = Area of the characteristic ion for the compound being measured.
A_{is} = Area of the characteristic ion for the specific internal standard.
C_{is} = Concentration of the specific internal standard.
C_X = Concentration of the compound being measured.

The RF is calculated for each standard concentration level in the initial calibration curve. The average RF is then calculated off these points.

Sample Calculations

Sample calculations are made by the instrument software and are performed as follows:

Water

$$\text{Concentration (ug/L)} = \frac{(Ax)\,(Is)}{(Ais)\,(RF)(V_o)}$$

where:

A_X = Area of characteristic ion for compound being measured.
Is = Amount of internal standard injected (ng).
A_{is} = Area of characteristic ion for the internal standard.
RF = Average response factor for compound being measured.
V_o = Volume of water purged (mL).

Soil: (low level)

$$\text{concentration (ug/Kg)} = \frac{(A_X)(Is)}{(A_{is})(RF)(W_g)}$$

where:

A_X = Area of characteristic ion for compound being measured.
Is = Amount of internal standard injected (ng).
A_{is} = Area of characteristic ion for the internal standard.
RF = Average response factor for compound being measured.
W_g = Weight of sample purged (g).

EXAMPLE CALCULATIONS

Sample Calculations

Sediment/Soil, Sludge, and Waste: (medium level)

$$\text{Concentration (µg/kg)} = \frac{(A_x)(Is)(V_t)}{(A_{is})(RF)(V_i)(W_g)}$$

where:

A_x, Is, A_{is}, RF = same as for soil above.
V_t = Volume of total extract (uL) (use 10,000 uL or a factor of this when dilutions are made).
V_i = Volume of extract added (uL) for purging.
W_g = Weight of sample extracted (g).

Instrument software is not set up to account for the medium level extraction procedure. Concentrations appearing on the quantitation report must be multiplied by a factor of 125 to correct for sample weight and extract volumes.

Sample Calculations

Sample No.: 006613-01MS Compound: 1,2,4-Trimethylbenzene

Matrix: Tissue

$$\text{Concentration (ug/kg)} = \frac{(606139)(250\,ng)(__)}{(252992)(3.351)(1g)(__)} = 179$$

Example Calculations Performed By: US Date: 3/20/14

Miscellaneous Additional Data
(If applicable)

☒ **None** 3/24/14

Rev. 1.1